Editorial

J.		Sontheimer
E.H.		Farthmann

The development of stenosis or respiratory fistula is a devastating complication of late-stage carcinoma of the esophagus or bronchus. The quality of life is seriously decreased by this complication. Life expectancies at this stage average about 120 days. During the last 20 years different types of plastic prostheses have been introduced in order to treat malignant strictures in cases of local nonresectability or general inoperability of the patient. In 1986, in a great series of over 2,000 patients, Tytgat reported a procedure-related mortality rate of 4% and a perforation rate of 5% caused by bouginage and insertion of the tube. In the follow-up there were more than 30% with bolus impaction. The problem of the traumatizing insertion of the plastic prosthesis and the need of general anesthesia have led to the development of self-expandible metal stents. These stents have the advantage of a better ratio between inner and outer diameter. They seem to have lower complication and mortality rates. Until today no prospective randomized studies have been done to evaluate the different metal stents available, but there are a few studies reporting the advantages of metal stents in comparison with plastic tubes.

At present the stents mostly used are the coated Gianturco Rösch Z Stent, the coated Streckerstent and the Wallstent. Especially the coated stents are advantageous to cover esophagorespiratory fistulae and to avoid tumor ingrowth. The problem of stent migration is dependent on the experience of the user. For this reason no more than two different types of stents should be used in an endoscopic unit.

If these principles are considered the reported primary technical success rate is 100%, the rate of severe complications is 1.5% and the rates of minor complications, such as stent occlusion, less than 20% during the follow-up.

The metal stents available at the present time for the treatment of malignant stenoses and esophagorespiratory fistulas can be inserted very safely under mild sedation and represent a distinct improvement in comparison with plastic prostheses.

The costs of the metal stent systems are far too high. There is an urgent need for a more realistic price formation.

KARGEH

E-Mail karger@karger.ch Fax + 41 61 306 12 34
<http://www.karger.ch>

© 1997 S. KargerAG, Basel