Paper

Urol. int. 1974;29:224

Clinical Experience with the Mentor Bladder Stimulator

Preliminary Report

D.C.	Daniel C.	Merril
C.J.	Christopher J.	Conway

Department of Surgery, Division of Urology, University of Minnesota Health Sciences Center, Minneapolis, Minn.

Key Words

Bladder
Electrical stimulation
Neurological

Abstract

A new bladder stimulator which limits current spread to adjacent pelvic floor structures has been tested in four patients with a variety of neurological lesions. The device successfully emptied the bladders of all patients tested. However, electrical stimulation produced a **spinal-cord-mediated** pelvic floor contraction in the patient who had an upper motor-neuron lesion. Voiding was successful in the latter patient following chemical rhizotomy.

Author's address: Dr. Daniel C. Merrill, Department of Surgery, Division of Urology, University of Minnesota Health Science Center, Minneapolis, Minn. (USA)

Literature

Caldwell, K. P. S.: The electrical control of Sphincter Incompetence. Lancet ii: 174–175 (1963). Halverstadt, D. B.: Electrical stimulation of the human bladder: 3 years later. J. Urol., Baltimore 106: 673–677 (1971). Merrill, D. C; Glover, E. G., and Conway, C. J.: Electronic bladder control. Med. Electron. Data 2: 72 (1971).