State of the Art in Platelet Function Testing

Beate E. Kehrel^a Martin F. Brodde^{a,b}

^a Department of Anesthesiology, Intensive Care and Pain Medicine, Experimental and Clinical Hemostasis, University of Münster,
^b OxProtect GmbH, Münster, Germany

Keywords

Platelet function testing · Platelet disorders ·
Antiplatelet drugs · Platelet concentrates

Summary

Platelets perform many functions in hemostasis but also in other areas of physiology and pathology. Therefore, it is obvious that many different function tests have been developed, each one conceived and standardized for a special purpose. This review will summarize the different fields in which platelet function testing is currently in use; diagnostics of patients with bleeding disorders, monitoring patients' response to anti-platelet therapy, monitoring in transfusion medicine (blood donors, platelet concentrates, and after transfusion), and monitoring in perioperative medicine to predict bleeding tendency. The second part of the review outlines different methods for platelet function testing, spanning bleeding time, and platelet counting as well as determining platelet adhesion, platelet secretion, platelet aggregation, platelet morphology, platelet signal transduction, platelet procoagulant activity, platelet apoptosis, platelet proteomics, and molecular biology.

Introduction

Platelets are known to have a large variety of functions. They are main regulators of hemostasis, where they contribute to the hemostatic plug, and in addition accelerate the coagulation system [1]. Platelets also interact with a large variety of cell types, such as monocytes, neutrophils, endothelial cells and smooth muscle cells, and contribute to the pathogenesis of atherosclerosis and vascular inflammation [2–4]. Platelets also contribute to neurodegenerative diseases [5] and other

inflammation-mediated pathologies such as arthritis, systemic lupus and sepsis [7] as well as acute lung injury, including transfusion-related acute lung injury (TRALI) [8]. In addition platelets are involved in cancer metastasis [9] and in innate immunity [10] as well as in adaptive immunity [11, 12]. Platelet interact with microorganisms [13] and with neutrophils to provide protection against invasion by pathogens by forming neutrophil extracellular traps (NETs) [14]. Such NETs are involved in infection-mediated thrombosis, like disseminated intravascular coagulation and deep-vein thrombosis [15].

Platelet-derived microparticles and exosomes not only participate in hemostasis but also play a role in thrombotic and inflammatory diseases as well as in cancer progression and metastasis [16]. Accurate measurement of platelet functions is critical for basic research on the role of platelets in physiology and pathology, is necessary for identifying patients with platelet dysfunction or potential hyperfunction, is becoming probably important for the monitoring of antiplatelet therapy, and is in addition valuable to determine platelet function in donors, concentrates, and after transfusion as well as in managing perioperative hemostasis.

Platelets are sensitive to manipulation, and are prone to artificial in vitro activation. Therefore testing of platelet function requires a high degree of experience as well as expertise to perform and interpret. Unexpected, abnormal platelet function tests should always be repeated with a fresh blood sample and in parallel with a normal control sample. As there are so many different platelet functions and so many different application areas, such a thing as we call in German 'eierlegende Wollmilchsau', an all in one method/device suitable for every purpose, to study platelet function cannot exist. For example, special devices for the assessment of anti-platelet drugs, for studying platelet function under shear conditions, or for testing physical properties of the clot have been developed. Such assays are developed and standardized for a special purpose. Using these function tests for other purposes might lead to misinterpretation or over-interpretation of the results and harm to patients.

Platelet Function Testing in Patients with Bleeding Disorders

An evaluation of patients with abnormal bleeding requires objective clinical assessment of bleeding history, family history, and physical examination. When appropriate, it is followed by laboratory investigations. A carefully collected history provides the most effective tool for determining presence and significance of a bleeding disorder. If bleeding symptoms cannot be fully explained by standard laboratory investigation of coagulation, fibrinolysis, or von Willebrand's disease, laboratory investigations of platelet number and function are recommended [17]. To help to determine the cause of or potential for excessive bleeding and management of patients with platelet disorders, the multi-step process to assess platelet function starts with platelet count and size. If appropriate, platelet adhesion studies, platelet aggregation testing, platelet secretion studies, and specific testing including flow cytometry, electron microscopy, study of signal transduction pathways, support of thrombin formation, genetics, and proteomics will be done [18–20].

In most cases a platelet-mediated hemostatic disorder cannot be characterized by just a single function defect, but rather by a combination of platelet functional abnormalities. Platelet function disorders are varied, with a wide spectrum of known disorders [21, 22], and therefore many different function tests are necessary for a clear diagnosis. This is a significant challenge even for specialized laboratories. The different function tests which are in use for the assessment of platelet function in patients with bleeding problems are described below.

A diagnostic approach to platelet disorders in children has recently been published [23]. Children are not 'small adults', and platelet function differs profoundly [24–26]. Therefore, 'general standard values', according to the age of the children, are inevitable to generate for each single platelet function test in each different laboratory [27]. Molecular genetic diagnosis of heritable platelet disorders may offer valuable confirmation of diagnosis in affected individuals, in family members and for ante-natal diagnosis, and is valuable for the understanding of platelet function. Single-nucleotide polymorphism microarrays (SNP arrays) are able to map regions of homozygosity and to narrow areas of interest for further directed sequencing in families with shared genetic heritage and with proven consanguinity. Massively parallel DNA sequencing technologies (next-generation sequencing) have rendered the whole-genome re-sequencing of individual patients practical [28, 29], but such analyses are currently only performed in research laboratories and are very expensive. This repertoire can be extended by 2D-DIGE platelet protein analysis (whole platelets, phospho-proteome and platelet secretome) and even for a few cases complete platelet proteomics [30].

Platelet Function Testing to Monitor Patient Response to Antiplatelet Therapy

Platelet responsiveness to anti-platelet drugs might carry direct clinical relevance, but the mechanisms leading to poor anti-platelet drug effects are not fully elucidated. Studies assessing the ability of platelet function tests to predict patients' clinical response to aspirin or clopidogrel have generated contradictory results. Therefore, the use of platelet function tests to monitor and guide anti-platelet therapy has provoked heated debates [31, 32].

On the one hand, Gurbel et al. [33] and Geisler et al. [34, 35] described that low response to clopidogrel in patients with symptomatic CAD treated by stenting significantly enhanced the occurrence of cardiovascular events and death. Several other authors share the positive view on the role of platelet function testing for making informed decisions on both initiation and titration of anti-platelet drug therapies, so-called tailored anti-platelet therapy. Sharma et al. [36] wrote recently: 'These platelet function assays are no longer regarded just as a laboratory phenomenon but rather as tools that have been shown to predict mortality in several clinical trials. It is believed that suboptimal response to an anti-platelet regimen (pharmacodynamic effect) may be associated with cardiovascular, cerebrovascular, and peripheral arterial events'.

On the other hand, a study on 2,440 patients, 1,227 in the monitoring group, reported recently [37] that patients, who received platelet function monitoring by the VerifyNow assay and dose adjustment after stent implantation, fared no better than those given standard care. Additional studies are needed to determine whether changes in therapy based on results of platelet function testing improve clinical outcomes. One of these studies will be the ongoing ANTARCTIC study which assesses the value of platelet function testing in patients aged >75 years with the focus on the prevention of major bleedings.

The absence of an association between stronger therapy with anti-platelet drugs and clinical outcomes has been described [38, 39]. In patients with a poor response to clopidogrel, intensification of platelet ADP receptor inhibition failed to improve outcomes [40, 41]. Therefore, the strategy of monitoring platelet function to yield better ischemic and safety outcomes might be still hampered by at least two uncertainty factors: What is the optimal platelet function test for this purpose, and what is the optimal therapy of identified high risk patients?

The prognostic value of high on-treatment platelet reactivity against ADP or arachidonic acid has been shown repeatedly, but that does not inevitably mean that inadequately inhibited ADP receptors or cyclooxygenase are the cause of patients' higher risk. Platelet response to ADP and/or arachidonic acid might just be a marker, and we still have to find out the cause of patients' higher risk. Therefore, other methods as the yet tested ones might be necessary to develop based on deeper understanding of pathology of thrombosis and ischemia. In addition, if high on-treatment platelet reactivity to ADP and/or ara-

chidonic acid is just a risk marker and not more, we might need to reconsider the proposed changes in treatment of these patients. This would mean that platelet function testing is still valuable, as we identify patients at higher risk, but we need to find new strategies to overcome these higher risks.

Several function tests have been used to monitor the presence and effectiveness of anti-platelet medications. Global assessments of 'hemostasis', such as the standard thrombelastograph, Sonoclot, Clot Signature Analyser and Hemodyne, are not specific for platelet function and are essentially insensitive to cyclooxygenase inhibitors and P2Y₁₂ antagonists [42]. Of the platelet function tests assessed by Breet et al. [43], only light transmittance aggregometry with ADP as the agonist, VerifyNow P2Y₁₂, and Plateletworks using ADP tubes, were significantly associated with the primary end point, and the predictive accuracy of these tests was only modest [43]. The VerifyNow assay, the best evaluated assay so far, is still limited by the poor positive predictive value (12.7%) [44].

Other methods assess activation-dependent changes on the platelet surface by flow cytometry (vide infra). These tests include measurement of levels of platelet surface P-selectin, CD63, activated GPIIb/IIIa, PAR-1 cleavage, microparticle formation and monocyte-platelet association. Flow cytometric measurement of platelet activation status and platelet reactivity to closely monitor anti-platelet drug therapy is been hampered by the fact that the stabilization of the blood sample was a real challenge. This problem might in the future be overcome by the use of specially developed fixatives (Platelet Solutions Ltd, Nottingham, UK) that stabilize blood samples so that measurements can be performed up to 9 days after blood sampling. This allows blood samples to be transferred by post to a central laboratory where the analyses are performed and negates the need for any special equipment or technical expertise at the point of blood sampling.

Platelet Function Testing in Transfusion Medicine

In transfusion medicine platelet function is mainly determined in blood donors, platelet concentrates, and after transfusion [45].

Platelet Function Methods to Screen Donors

Platelet concentrates can be prepared as random-donor platelet concentrates from whole blood or as apheresis platelets from a single donor. Single donor apheresis concentrates more and more account for a large percentage of platelet concentrates. Therefore, the platelet quality of individual donations becomes increasingly important. The best 'function test' is, as mentioned in tests for screening of platelet disorders, a good anamnesis, including drug history. Platelet function screening can identify donors with diet-related platelet dysfunction or with poor recollection of aspirin use [46–48]. Platelet refractoriness induced by alloimmunization can be

managed by transfusion of matched platelets [49]. Assays like microarray-based genotyping or flow cytometric cross-match can be used for this purpose [50, 51].

Assessment of Platelet Function and Activation Status in Platelet Concentrates

The main application of platelet tests in transfusion medicine is to determine platelet function in concentrates. To produce platelet concentrates, platelets are mechanically manipulated, removed from the physiological environment, and stored. Therefore, it is important to evaluate platelets' hemostatic properties as well as changes, like platelet activation or platelet apoptosis, that might be an indicator for shortened survival after transfusion [52] or even 'prognostic indicators' for possible transfusion complications as longer storage of buffy coat-derived platelets was associated with inflammation and an increased risk of TRALI [53–56].

Platelet's hemostatic capability is often tested by application of the low-shear methods (light transmission aggregometry, impedance aggregometry, multiple electrode aggregometry, flow cytometry), and methods applying high shear (the platelet function analyzer-100 (PFA-100) and the cone and plate(let) analyzer). An overview on platelet function tests in stored concentrates is given in recent reviews [45, 57, 58].

Not everything that can be measured needs to be measured. In validation studies it is more important to get knowledge on the influence of changes in the procedure to gain and store platelet concentrates on the platelets as it is in the routine ongoing check of platelet concentrate quality. During platelet storage extracellular plasma proteins inside the concentrates are subjected to aging and other influences that might change their conformation and function. As such proteins influence hemostasis and inflammation, it might prove beneficial to quantify such altered proteins in stored platelet concentrates. This can be done by direct quantification using tools that recognize the altered protein structures or by platelet function tests using gel-filtered fresh donor platelets and plasma from concentrates as such proteins activate platelets (personal communication by the authors).

Testing of platelet activation status in concentrates is mostly done by observing platelet 'swirling', the shimmering of discoid platelets, and measuring ristocetin-induced aggregation (GPIIb-related function), lactate dehydrogenase (LDH) concentration, pH value, platelet factor 4 (PF4) release, and cell surface expression of activation markers like P-selectin [59, 60]. Measuring platelet apoptosis (vide infra) and/or microparticle formation (vide infra) in platelet concentrates are new emerging fields [61]. Understanding of platelet apoptosis and its role in the platelet storage lesion is an exciting challenge [62].

Assessment of Platelet Function Following Platelet Transfusion

To measure the outcome of platelet transfusion is still an unsolved challenge. Quantifying bleeding tendency as an outcome measure seems to have the highest clinical relevance,

but key variables like severity and site of bleeding and differences and methods trying to quantify bleeding make this approach not really reliable for studies [63]. Therefore, in most cases platelet recovery and survival is measured as a surrogate parameter to assess the clinical function of platelet concentrates. In vitro evaluation of stored platelets did not fully predict post-transfusion platelet survival and function [64]. Platelet recovery can be measured by radiolabelling an aliquot of platelet concentrate and re-infusion into the normal volunteer donor. Platelet recovery can also be measured in animals using flow cytometry and anti-human platelet antibodies [65]. These tests have been helpful to study the clearance mechanism of chilled platelets [66, 67].

Perioperative Assessment of Platelet Function

Whether objective measure of platelet function may aid in perioperative hemostatic control is still a matter of debate. It is unclear so far whether platelet testing in patients with no bleeding history adds to the care of patients at all and, if though, which patients may actually benefit from routine pre- and perioperative testing of platelet function. Several studies are in favor of perioperative platelet function testing. Poston et al. [68] for example described platelet aggregometry to be useful to predict bleeding and thrombosis after off-pump coronary artery bypass. Other authors described that preoperative screening does not add to patients' care and might even result in unnecessary medication [69, 70]. Current evidence does not suggest the use of platelet function tests to guide therapy with anti-platelet drugs in the perioperative period [71].

The association between on-treatment platelet reactivity measured by an ex vivo assay and the occurrence of bleeding events is low [72]. The bleeding time has been widely used as a preoperative screening test, but it is a completely non-informative test, and therefore not beneficial for this purpose [73, 74]. Carroll et al. [75] came to the conclusion that the most significant predictors of postoperative bleeding were a low body mass index (BMI), lowest core body temperature, and cross clamp time, but not perioperative platelet function tests (glass bead adhesion and thrombelastograph platelet mapping). Korte et al. [76] used preoperative fibrin monomer measurement for risk stratification for high intraoperative blood loss in elective surgery. Using the Enzymun Test FM, preoperative fibrin monomer concentration $< 3 \mu\text{g/l}$ excluded intraoperative blood loss $> 500 \text{ ml}$ with 92% sensitivity and 95% negative predictive value, while platelet count did not. Using other fibrin monomer assays, this remarkable result could not be repeated. The difference in methodology is not yet understood. In conclusion the authors agree with Peterson et al. [77]: 'The best preoperative screen to predict bleeding continues to be a carefully conducted clinical history that includes family and previous dental, obstetric, surgical, traumatic injury, transfusion, and drug histories'.

Methods to Test Platelet Function

Template Bleeding Time

The bleeding time is a crude test of hemostasis. Bleeding time, the time taken for bleeding to stop after a defined incision is made into the skin, to measure the interaction between platelets and blood vessels to form a clot has been developed by Duke in 1910 [78]. Ivy [79] made the method more reliable, by introducing a blood pressure cuff on the upper arm, inflation to 40 mm Hg and placing the incision into the anterior surface of forearm. Spring-loaded template devices are provided by different companies to make standard sized cuts within the skin.

Bleeding time has been used most often to detect qualitative defects of platelets, vascular defects, or von Willebrand's disease. It is also prolonged in thrombocytopenia; therefore, it is important to check the platelet count before performing a bleeding time. Although it seems to be simple, bleeding time measurement is invasive, poorly reproducible, and relatively insensitive to platelet function defects. Therefore, it is of limited aid in evaluating individual patients and useless in prediction of the risk of abnormal bleeding from internal organs during and/or after an invasive procedure [73, 74], although is still in use in some centers.

Lehman et al. [80] wrote in their recently published article, showing that discontinuation of bleeding time tests at their University had no detectable adverse clinical impact: 'Unfortunately, the BT test became popularized in clinical practice without sufficient peer-reviewed evidence supporting its use, probably under the assumption that an in vivo test was needed and no other test was available'.

Platelet Counting Methods

Rapid and accurate platelet concentration measurements are crucial in all laboratories working with platelets. Platelet counting is clinically required in severe thrombocytopenia. Several techniques have been developed to count platelets. The manual count is still recognized as the gold standard or reference method especially for measurement of low platelet counts, but it is time-consuming and requires a high degree of technical skill [81]. Automated hematology analyzers can be used to obtain platelet counts over a wide range of values; but cell size analysis cannot discriminate platelets from other similar-sized particles like small erythrocytes or erythrocyte fragments. Optical counting methods like light scatter or fluorescence methods have been introduced for automated platelet counting. The RBC/Platelet Ratio International Reference Method (IRM) uses flow cytometry to count large numbers of cells rapidly, accurately, and precisely [82, 83]. Platelet count is derived from the ratio of fluorescent platelets, labeled with a monoclonal antibody, to collected red blood cells. Recently a flow cytometer, the compact BD Accuri™ C6, has been developed that can, due to a unique fluidic system driven by peristaltic pumps, report absolute cell counts based on direct

volume sampling. This new direct volume method requires no additional measurement of red blood cells or reference beads and therefore reduces costs and measurement time [84].

Platelet Adhesion and Spreading

The adhesive properties of platelets contribute to maintaining vascular integrity and defending from external pathogens [85]. Platelets patrol the blood vessels and adhere where alterations of the endothelial cell lining are detected. Alterations of the endothelium can be accompanied or not by exposure of subendothelial matrix components. Adhesion to the exposed subendothelial extracellular matrix (ECM) proteins is the first step in the formation of a hemostatic plug. In addition the adhesive property of platelets to foreign surfaces such as implants or catheters can be measured.

The different platelet adhesion assays can be divided in those which measure platelet adhesion performed under static conditions and those which measure platelet adhesion under flow conditions. A nice overview about platelet adhesion assays under static conditions is given by Joanne Stevens [86].

Platelet adhesion can be measured in vitro using cultured vascular cells [87] or foreign surfaces such as glass beads, but today mainly purified matrix proteins, like collagen, von Willebrand factor, thrombospondin [88] and fibrinogen, or complete ECM from cultured endothelial cells are used. Platelet adhesion can also be evaluated on cryostat cross-sections of human coronary arteries with or without atherosclerotic plaques [89] or on cholesterol sulphate [90]. Adhesion to glass beads has been introduced by Hellem [91] in 1960 and Salzman [92] in 1963. Today this method is out of fashion. Perfusion chambers with matrix proteins or complete ECM have been introduced in the 1970s [93–95]. Under flow conditions adherence of platelets is affected by rheological conditions such as shear rate, presence of red blood cells, red blood cell deformability, and viscosity of the medium [96, 97]. Later perfusion chambers were adapted to smaller volumes of blood [98]. New improved flow systems led to the introduction of blood flow pulsatility [99]. Nowadays techniques with a variety of microfluidic devices are in use to observe platelet adhesion. Biochips, like the Cellix Vena8 Fluoro-TM biochip, can be coated with collagen and platelet adhesion can be measured using fluorescence immunostaining and confocal microscopy at up to 60 dynes/cm². Such chips can contain several parallel microcapillaries for continuous flow that can be coated with different adhesion molecules. Down-scaling of the blood volume now allows tests to be done more quickly and with better standardization [100–102].

Spreading

Platelet spreading depends on platelet cytoskeletal rearrangement. The methods which are used today most frequently are fluorescence microscopy, using marker-conjugated phalloidin, or scanning electron microscopy [103, 104].

Mixed Adhesion/Aggregation and Shear Conditions

Several in vitro methods measure platelet adhesion together with aggregation under shear conditions. Among these methods are the platelet function analyzer (PFA 100®) and the Impact-R® system (Cone and Plate(let) Analyzer; CPA). A review on the CPA is given by Savion and Varon [105] in 2006. The method is very sensitive for von Willebrand factor-mediated platelet activation processes and has been described to allow differentiation between inherited and acquired thrombotic thrombocytopenic purpura [106].

The PFA 100 test is a rapid, accurate test of monitoring shear-dependent platelet function, as anticoagulated blood is passed through a membrane coated either with collagen plus epinephrine (Col/Epi) or collagen plus ADP (Col/ADP) at a high shear rate. Platelets adhere to the membranes and gradually occlude a small aperture in the center of each membrane. The time to occlude the aperture in the membrane is the so-called closure time. As for all shear-dependent methods, knowledge of the full blood count is critical for interpreting results from the PFA-100 and Impact R. Abnormal platelet adhesion using both methods is typical for types 2A, 2B, 2M, and 3 VWD [107]. A normal PFA 100 closure time result should not be used to rule out platelet function defects [108]. The PFA-100 does not predict delta-granule platelet storage pool deficiencies [109].

Intravital Imaging of Thrombus Formation

Recently intravital imaging of thrombus formation has been reviewed by Kuijpers and Heemskerk [110] as well as by Furie and Furie [111]. As these function tests are restricted to animals, they are not discussed in this article.

Measurement of Total and Released Nucleotides, Released Serotonin, and Released PF4

Measurement of total and released nucleotides (ATP, ADP) is used for determining specific deficiencies in the number of dense granules and in dense granular content as in the so-called storage pool defects or specific defects in degranulation as in release defects [112].

In the past it has most often been done in combination with aggregometry using a lumi-aggregometer, but this method does not allow distinguishing between storage pool and release defects. Therefore, in several laboratories the total amount of ADP and ATP in lysates of resting platelets and in platelet releasates is measured in addition. The methods for ATP/ADP measurement are either based on luminescence [113] or on HPLC [114]. ADP is measured as ATP after phosphorylation by pyruvate kinase (PK) with phosphoenolpyruvate (PEP).

Serotonin (5-HT) is actively taken up and stored within the platelet dense granules. It is measured for example at suspicion of storage deficiency like in Hermansky-Pudlak Syndrome (HPS) platelets or storage pool disease. 5-HT uptake and release has been measured using radiolabelled 5-HT [115]. This assay has specially been used for heparin-induced

thrombocytopenia II (HIT II) diagnostics [116]. Platelet 5-HT content today is also measured by ELISA HPLC or by a fluorescence-based assay using a fluorescence microscope or a flow cytometer [117]. Recently, a technique capable of measuring real-time 5-HT release has been described based on fast scan cyclic voltammetry using carbon-fiber microelectrodes [118]. Platelet content and release of alpha granule proteins (e.g. PF4, beta-thromboglobulin and thrombospondin-1) can be measured by commercially available ELISAs.

Platelet Aggregometry

Aggregometry and flow cytometry are fundamentals in platelet function testing today. Light transmission aggregometry has been developed by the group of Gustav Born. Born's paper [119] is one of the most cited among platelet researchers. In light transmission aggregometry platelet rich plasma (PRP) is stirred in a cuvette which sits between a light source and a photocell. Agonists induce platelet adhesion to platelets and light transmission increases. Platelets will only aggregate if they are in contact with each other. So they must be stirred while testing is taking place. In routine practice ADP, collagen, ristocetin, arachidonic acid, adrenaline, and PAR-1-activating peptide (TRAP-6, SFLRN) are used. But the panel of agonists can be extended. Other agonists are for example gamma thrombin or alpha thrombin (in the presence of GPRP to prevent fibrin polymerization), TRAP-4 (AYPGKF), thromboxane mimetic U46619 (stable analogue of the endoperoxide prostaglandin H₂) [120], calcium ionophore A23187, polymerized immunoglobulins, collagen-related peptide (CRP) [121], misfolded proteins like AGE-proteins and oxLDL [122], alpha defensins [123], proteins from microorganisms like EAP [124], and snake venoms like convulxin [125]. The aggregation parameters to consider are the lag phase, maximal amplitude, primary aggregation slope, and disaggregation for each used agonist concentration [126, 127]. Several types of mild platelet defects, (e.g. primary secretion defects) may be missed by aggregation. Reversible first-wave aggregation is seen with ADP and adrenaline and suggests a failure of granule release as in platelet storage pool disorder or a defect in nucleotide release [128, 129]. A lack of agglutination with ristocetin is consistent with the possible diagnoses von Willebrand disease or Bernard Soulier syndrome [130], while possible diagnoses with no aggregation with ADP, adrenaline, TRAP-1, and collagen are consistent with Glanzmann's thrombasthenia or afibrinogenemia.

Laser light scatter aggregometer are especially suitable to measure the formation of small aggregates [131]. Spontaneous platelet aggregation, defined as formation of small aggregates under constant stirring in the absence of any agonists, is rare in healthy individuals but seen in some cases of von Willebrand disease [132], in some patients with diabetes [133] and in some lipid disorders [134], and may be an indicator of an ongoing active atherosclerotic process [135].

Impedance aggregometry, developed in the early 1980s [136], allows the use of whole blood which is stirred at 37 °C; aggregation is detected by the adhesion of platelets to the surface of two fine, metal wire electrodes [137]. The electrical impedance between the electrodes can be displayed as a wave of aggregation. Although the use of whole blood makes impedance aggregometry highly valuable as it measures platelet function within whole blood without the requirement of sample processing, there are several drawbacks of impedance aggregometry in comparison to light transmission aggregometry that researchers and clinicians need to have in mind. The reversibility and a biphasic response to ADP cannot be demonstrated by whole blood impedance [138, 139], and platelet response to epinephrine/adrenaline tends to be absent or very weak [140]. The development of disposable electrodes, standardized reagents, and the availability of a 5-channel multiple electrode platelet aggregometer has solved several standardization problems [141].

The Plateletworks methodology involves using a cell counter to measure total platelet count in a whole blood sample and then to re-determine the platelet count on a second sample that has been exposed to a known platelet agonist. Aggregated or agglutinated platelets will not be counted as platelets in the second sample. The test has especially been used for assessment of anti-platelet drugs [142].

Another methodology, the VerifyNow assay that has been developed for monitoring of anti-platelet drugs in whole blood is based on platelet aggregation [143]. The VerifyNow P2Y₁₂ assay evaluates platelet aggregation of fibrinogen-coated beads in response to ADP and prostaglandin E₁. Light transmittance increases as activated platelets bind and aggregate fibrinogen-coated beads. The cut-off value for defining suboptimal response to ADP-platelet aggregation despite treatment with clopidogrel has not been firmly established [144]. The VerifyNow Aspirin assay incorporates the agonist arachidonic acid to activate platelets, while in the VerifyNow IIB/IIIa assay the reagent iso-TRAP is incorporated to induce platelet activation without fibrin formation.

Flow Cytometric Analysis of Platelet Functional Capability and Activation Markers

The ex vivo or in vitro testing of functional platelet properties using conventional techniques reflects the overall behavior of the whole platelet population in the sample under investigation. Flow cytometry allows the analysis of individual platelet functional capability and the measurement of the expression of platelet activation markers on individual platelets as well as the quantitation of associates between platelets and other blood cells [145, 146]. Overviews on the methodology of platelet flow cytometry are given by Michelson [147, 148] and Knight [149].

A step-by-step instructions manual to perform platelet flow cytometry written by Brodde et al. in 2012 is given in a booklet provided by Becton Dickinson [150]. It includes instructions

and samples of results from patients with inherited thrombocytopathies and examples of results for monitoring of anti-platelet drugs. Flow cytometry allows detecting platelet function defects even in patients with very low platelet count and therefore is the methodology of choice to determine whether a thrombopenic patient has in addition a thrombocytopathy.

Platelet procoagulant activity can be measured by flow cytometry at different steps, by the expression of phosphatidylserine on activated platelets using labeled annexin V [151], and by the binding of coagulation factors using directly labeled coagulation factors or the combination of factor and specific antibody [152–154]. The determination of platelet-specific autoantibodies by flow cytometry has been reviewed by Tomer [155]. As diagnostic aid in the evaluation of thrombocytopenic disorders a flow cytometric analysis of thiazole orange to measure reticulated platelets has been developed [156]. Flow cytometry also allows to measure steps in signal transduction as intracellular methods have been developed. The most often used method is the intracellular quantification of VASP phosphorylation [157]. In addition flow cytometry is the method of choice to measure platelet microparticles as well as microparticles from other cells [158].

Quantitation of Shedded Platelet Receptors

Metalloproteinase-mediated ectodomain shedding of platelet receptors has emerged as a new mechanism for modulating platelet function [159]. Interestingly 5-HT stimulates platelet receptor shedding by ADAM17 [160]. Quantitation of shedded soluble platelet receptors in plasma has been used to analyze platelet function in different diseases. The first protein known to be shedded from the platelet surface was GPIb. The soluble shedded part has been named glycocalicin [161]. Quantitation of glycocalicin has especially been used in thrombocytopenic patients [162]. Recently it was described that diesel-exhaust particles are associated with GPIb shedding [163]. Shedding of CD36 from platelets and other blood cells has been associated with type II diabetes [164]. Plasma levels of shedded platelet glycoprotein V have been linked to type 2 diabetes, stroke, and myocardial infarction [165–167]. Shedded GPVI is evaluated as marker for acute coronary syndrome [168], while soluble CD40L, an immunostimulatory ligand for CD40 released from platelets, is associated with adverse reaction to platelet transfusion [169].

Signaling Pathways

Many different methods have been developed for the evaluation of platelet signal transduction. A series of reviews has been published recently for these methods. Nesbitt et al. [170] described a live cell microimaging technique to examine platelet calcium signaling dynamics under blood flow. Measurement and manipulation of Ca²⁺ in suspensions of platelets as well as in single platelets and megakaryocytes has been outlined by Ohlmann et al. [171] and Mason et al. [172], respectively. An up-to-date overview on signaling receptors on

platelets and megakaryocytes is given by Woulfe et al. [173]. How to study protein phosphorylation and dephosphorylation is presented by Gibbens [174], by Jones and Poole [175], and by Gorter and Akkerman [176]. Shrimpton et al. [177] described the isolation and analysis of platelet lipid rafts. The evaluation of redox-signaling is described by Naseem et al. [178]. Receptor tyrosine kinases (RTKs) and G-protein-coupled receptors (GPCRs) can form platforms in which protein signaling components specific for each receptor are shared to produce an integrated response upon engagement of ligands [179]. Evaluation of these signaling platforms allows a different view on platelet signaling in comparison with the model of cross-talk between RTKs and GPCRs.

In several techniques to study platelet signaling megakaryocytes are used as surrogate for platelets. These techniques include patch clamp, interference RNA, quantitative PCR, Ca²⁺-sensitive fluorescent indicators, and transgenic murine models. Experimental and computational methods can be combined to study platelet signaling [180]. A good example is the PlateletWeb [181, 182].

Platelet Proteomics

Initial platelet proteomic research focused on general proteome mapping. Exploration of subcellular compartments, the membrane proteome, and signaling pathways followed. Proteomics technology is based on mass spectrometry in combination with several separation techniques such as two-dimensional gel electrophoresis or multidimensional liquid chromatography. Differential proteomics can be used as a discovery tool to identify functionally important proteins in platelet activation as in platelets several proteins change in abundance upon platelet activation. This technique helps to identify activation networks. Two-dimensional difference gel electrophoresis (2D-DIGE) is a brilliant method to overcome the problem of inter-gel variability, connected with classical two-dimensional gel electrophoresis and therefore is an ideal discovery tool [183]. These techniques allow also the comparison of platelet proteome between different patients and between patient groups [184, 185].

The strategy to combine 2D-DIGE with matrix-assisted laser desorption/ionization time-of-flight (MALDI-TOF) mass spectrometer analysis has been a useful approach for a quantitative analysis of the effect of thrombin on the secretome profile of human platelets [186]. To identify platelet surface proteins, these can be enriched using lectin affinity chromatography, biotin/NeutrAvidin affinity chromatography and free flow electrophoresis prior to identification [187]. Proteome analysis is useful in deciphering signaling cascades in human platelets [188, 189]. Combination of techniques to identify posttranslational modifications with techniques of proteomics led to the identification of palmitoylated platelet proteins and platelet phospho-proteome [190]. Platelet proteomics can also be applied to study pathologies where platelets play a fundamental role, e.g., acute coronary syndrome,

neurodegeneration, and diabetes [191–193]. Last but not least, platelet proteomic techniques are useful to study platelets with functional defects in more detail [194, 195].

Genetic Testing

Due to their enucleate structure, platelets are not accessible to large-scale genomic screens, but genetic testing is frequently used in patients with inherited platelet disorders (see above). The techniques used are above the scope of this review, and the reader is referred to the excellent review of Nurden et al. [196].

Measurement of Platelet Apoptosis

Similar to nucleated cells, platelets may undergo suicidal death, which is characterized by cell shrinkage, cell membrane blebbing, microparticle formation, cell membrane phospholipid scrambling, activation of caspases, cytochrome C release, and depolarization of the mitochondrial transmembrane potential ($\Delta\Psi_m$) [197]. Platelets have a functional intrinsic apoptotic signaling pathway. They contain the pro-apoptotic mitochondrial protease Omi/HtrA2 and Smac/Diablo as well as their target the X-linked inhibitor of apoptosis XIAP [198]. Detection of markers of platelet apoptosis is reviewed by Gylkhandanyan et al. [61]. In transfusion medicine, platelet apoptosis is a promising new field as platelets undergo apoptosis under nonoptimal storage conditions [199–201]. Platelet apoptosis can be provoked by suppression of thrombopoiesis, by injection of tumor necrosis factor, by high concentrations of thrombin [202], by anti-platelet antibodies and by contact with high concentrations of misfolded proteins (Brodde, Kehrel unpublished) and is seen in malaria infection.

Methods commonly used to study platelet apoptosis markers include flow cytometry, Western blot analysis, and electron microscopy. Depolarization of $\Delta\Psi_m$, an important parameter of mitochondrial function, can be measured by flow cytometry. The cytofluorimetric, lipophilic cationic dye, 5,5',6,6'-tetrachloro-1,1',3,3'-tetraethylbenzimidazolyliumcyanine iodide (JC-1) can selectively enter into mitochondria and change color from green fluorescence to intense red fluorescence as the membrane potential increases. In apoptotic platelets, JC-1 stays monomeric with green fluorescence [203].

Platelet-Mediated Thrombin Generation

Platelets play a major role in localizing and controlling the burst of thrombin generation leading to fibrin clot formation [204]. Fibrin at the vessel wall is built by the extrinsic coagulation pathway, but fibrin inside the body of the thrombus and fibrin in thrombus tails that form on the downstream margin of the thrombus is believed to be primarily dependent on the procoagulant function of platelets [205]. Platelet procoagulant function is tightly linked to two distinct processes: the induction of platelet activation and the induction of platelet cell death pathways, apoptosis, and necrosis. Two distinct pathways regulate platelet phosphatidylserine (PS) exposure and

procoagulant function [206]. In addition to platelets also red blood cells actively contribute to thrombin generation in whole blood [207]. Procoagulant tenase and prothrombinase complexes assemble on activated platelets to support cleavage of FX and prothrombin into the proteolytically active forms FXa and thrombin. In addition, platelets provide a scaffold for the formation of fibrin fibers and regulate clot retraction [208–210]. The GPIb thrombin binding site is essential for thrombin-induced platelet procoagulant activity [153]. Three steps in platelet procoagulant activity are usually measured: the exposure of surface PS, the binding of coagulation factors, and thrombin generation parameters. PS exposure is most often measured using labeled annexin V. This can be done fluorimetrically or by a flow cytometer [211]. By exposure of PS, platelets display high-affinity binding sites for vitamin K-dependent coagulation factors and coagulation inhibitors (protein C, protein S). Protein S may counterbalance platelets' procoagulant activity [212].

Also the binding of coagulation factors to activated platelets can be measured by flow cytometry as described by Doermann et al. [152]. Platelet factor V is sufficient to ensure thrombin generation in patients with severe inherited plasma factor V deficiency [213]. Feedback activation of thrombin generation on the platelet surface does depend on the activity of platelet disulfide isomerase (PDI) as surface-associated PDI is an important regulator of coagulation factor ligation to thrombin-stimulated platelets [154]. Not all activated platelets become procoagulant as seen by flow cytometry. Using collagen and thrombin (coat) as platelet agonist, the procoagulant platelet population can be increased. Dale et al. [214] showed that this procoagulant population, called 'coated platelets' use 5-HT to enhance the retention of procoagulant proteins on the cell surface. In addition to 5-HT, 'coated platelets' retain on their surface high levels of several procoagulant proteins, including fibrinogen, von Willebrand factor, fibronectin, amyloid beta, factor V, and thrombospondin that can be measured by flow cytometry.

Dale et al. [215] developed an easy-to-use method to identify 'coated platelets' by flow cytometry. PRP is activated by convulxin (a specific agonist for collagen receptor GPVI) plus thrombin in the presence of GPRP to inhibit fibrin polymerization and FITC-abciximab and biotin-fibrinogen are added. Biotin is recognized by phycoerythrin-streptavidin. The percentage of abciximab-positive platelets with bound biotin-fibrinogen is quantitated by flow cytometry. 'Coated platelets' are interesting biomarkers for prothrombotic diseases and in neurodegeneration [216–220].

The Thromboscope (CAT assay) (Thromboscope BV, Maastricht, The Netherlands) allows to measure parameters of thrombin generation in the presence of platelets (lag-time, maximum concentration of thrombin (Cmax), time required to reach Cmax (Tmax), and endogenous thrombin potential (ETP)). In the Thrombogram-Thromboscope assay thrombin measurement is carried out in a 96-well plate fluorometer. A thrombin

calibrator can be added to plasma. This calibrator is measured in the Thromboscope and compared to the measurement of fluorescence in another well containing the same plasma.

The replacement of the chromogenic substrate with fluorogenic substrates made it easier to study the role of platelets in coagulation. All platelet inhibitors tested thus far appear to influence thrombin generation in PRP. The test system is based on work by Hemker et al. [221, 222], who introduced the EPT as a term for the total amount of thrombin generated during the test.

Interestingly extracellular histones from dying cells and polyphosphates promote thrombin generation through platelet-dependent mechanisms [223]. Using assays for platelet-dependent thrombin formation might be useful to recognize new mechanisms in pathology and to find new avenues for anti-thrombotic drugs.

Assessing Protein Synthesis by Platelets

Platelets possess a translational machinery that can direct protein synthesis. A variety of protocols that can be employed to assess protein synthesis by platelets has been summarized by Schwartz et al. [224] in 2011. The interested reader is referred to this comprehensive instruction manual.

Measuring Platelet Contractile Force

Physical forces play a critical role in hemostasis by regulating the mechanobiology of platelets. Platelet contractions are driving the retraction and stiffening of clots and contribute in strengthening platelet adhesions through integrin-related mechanotransduction. The archetypical platelet disorder Glanzmann's thrombasthenia got its name from the Greek words 'thrombos' (thrombus, clot) and 'astheneia' (weakness) as clot retraction is disturbed in this disease. According to the review of Feghhi and Sniadecki [225], five different conventional platelet force assays have been developed, the traditional clot retraction assays, the in vitro observation of thrombus consolidation, the thrombelastography/-metry, the clot strip assay, and the clot retractometry [225–227]. Today techniques exist to measure even the physical forces of single individual platelets [228, 229]. Upon activation, G-actin monomers in platelets polymerize into F-actin filaments. This step

which is crucial for platelet mechanobiology can be measured by several assays. In platelet suspensions like PRP the DNase I inhibition assay can be used [230]. A method to measure platelet content of F-actin using rhodamine-phalloidin binding has been described by Cassimeris et al. [231]. A pyrenyl-actin assay of actin nucleation has also been described [232]. Today platelet cytoskeletal rearrangement is mostly analyzed flow cytometrically by measuring F-actin content with NBD- or bodipy-phalloidin or FITC-phalloidin [233, 234].

Measuring Circulating Platelet-Derived Microparticles

Circulating platelet – and other cell-derived microparticles have been implicated in several disease processes. Elevated levels are found in many pathological conditions. Flow cytometry is considered the 'gold' standard of microparticle detection [235], but also commercially available assays on 96-well plates that based on the procoagulant activity are in use [236]. Preanalytical parameters have a high impact on the measurement of microparticles [237, 238].

Morphological Methods

Electron microscopy has proven to be very useful for defining ultrastructural abnormalities associated with platelet defects and in basic understanding of platelet physiology [239–241]. A technical review is given by Clauser and Cramer-Borde [242]. Cheaper and compact table electron microscopes are currently developed.

As new techniques arrive at the horizon, they are also applied to platelets. Atomic force microscopy is one of these techniques. This technique has been already used, e.g., to study platelet interaction with biomaterials [243], for validation of flow cytometric detection of platelet microparticles [244], and even for the analysis of platelet GPIIb/IIIa receptor [245]. Differential interference contrast microscopy, also named Nomarski interference contrast microscopy, is an optical microscopy technique used to enhance the contrast in unstained, transparent samples. It has been used in several platelet laboratories with great success [246].

Disclosure Statement

The authors declared no conflict of interest.

References

- 1 Versteeg HH, Heemskerk JW, Levi M, Reitsma PH: New fundamentals in hemostasis. *Physiol Rev* 2013;93:327–358.
- 2 Jurk K, Kehrel BE: Pathophysiology and biochemistry of platelets. *Internist* 2010;51:1086, 1088–1092, 1094.
- 3 Nofer JR, Brodde MF, Kehrel BE: High-density lipoproteins, platelets and the pathogenesis of atherosclerosis. *Clin Exp Pharmacol Physiol* 2010;37:726–735.
- 4 Projahn D, Koenen RR: Platelets: key players in vascular inflammation. *J Leukoc Biol* 2012;92: 1167–1175.
- 5 Behari M, Shrivastava M: Role of platelets in neurodegenerative diseases: a universal pathophysiology. *Int J Neurosci* 2013; doi:10.3109/00207454.2012.751534.
- 6 Boilard E, Blanco P, Nigrovic PA: Platelets: active players in the pathogenesis of arthritis and SLE. *Nat Rev Rheumatol* 2012;8:534–542.
- 7 Katz JN, Kolappa KP, Becker RC: Beyond thrombosis: the versatile platelet in critical illness. *Chest* 2011;139:658–668.
- 8 Caudrillier A, Kessenbrock K, Gilliss BM, Nguyen JX, Marques MB, Monestier M, Toy P, Werb Z, Looney MR: Platelets induce neutrophil extracellular traps in transfusion-related acute lung injury. *J Clin Invest* 2012;122:2661–2671.

- 9 Labelle M, Hynes RO: The initial hours of metastasis: the importance of cooperative host-tumor cell interactions during hematogenous dissemination. *Cancer Discov* 2012;2:1091–1099.
- 10 Jurk K, Kehrel B: Platelets at the interface between hemostasis and innate immunity. *Transfus Med Hemother* 2004;31:379–386.
- 11 Semple JW, Freedman J: Platelets and innate immunity. *Cell Mol Life Sci* 2010;67:499–511.
- 12 Vieira-de-Abreu A, Campbell RA, Weyrich AS, Zimmerman GA: Platelets: versatile effector cells in hemostasis, inflammation, and the immune continuum. *Semin Immunopathol* 2012;34:5–30.
- 13 Niemann S, Spehr N, Van Aken H, Morgenstern E, Peters G, Herrmann M, Kehrel BE: Soluble fibrin is the main mediator of *Staphylococcus aureus* adhesion to platelets. *Circulation* 2004;110:193–200.
- 14 Ma AC, Kubes P: Platelets, neutrophils, and neutrophil extracellular traps (NETs) in sepsis. *J Thromb Haemost* 2008;6:415–420.
- 15 Fuchs TA, Brill A, Wagner DD: Neutrophil extracellular trap (NET) impact on deep vein thrombosis. *Arterioscler Thromb Vasc Biol* 2012;32:1777–1783.
- 16 Aatonen M, Grönholm M, Siljander PR: Platelet-derived microvesicles: multitasked participants in intercellular communication. *Semin Thromb Hemost* 2012;38:102–113.
- 17 Harrison P, Mackie I, Mumford A, Briggs C, Liesner R, Winter M, Machin S, British Committee for Standards in Haematology: Guidelines for the laboratory investigation of heritable disorders of platelet function. *Br J Haematol* 2011;155:30–44.
- 18 Hayward CP: Diagnostic approach to platelet function disorders. *Transfus Apher Sci* 2008;38:65–76.
- 19 Harrison P, Mumford A: Screening tests of platelet function: update on their appropriate uses for diagnostic testing. *Semin Thromb Hemost* 2009;35:150–157.
- 20 Zhou L, Schmaier AH: Platelet aggregation testing in platelet-rich plasma: description of procedures with the aim to develop standards in the field. *Am J Clin Pathol* 2005;123:172–183.
- 21 Jurk K, Kehrel BE: Inherited and acquired disorders of platelet function. *Transfus Med Hemother* 2007;34:6–19.
- 22 Nurden A, Nurden P: Advances in our understanding of the molecular basis of disorders of platelet function. *J Thromb Haemost* 2011;9(suppl 1):76–91.
- 23 Israels SJ, Kahr WH, Blanchette VS, Luban NL, Rivard GE, Rand ML: Platelet disorders in children: a diagnostic approach. *Pediatr Blood Cancer* 2011;56:975–983.
- 24 Mull MM, Hathaway WE: Altered platelet function in newborns. *Pediatr Res* 1970;4:229–237.
- 25 Israels SJ, Rand ML, Michelson AD: Neonatal platelet function. *Semin Thromb Hemost* 2003;29:363–372.
- 26 Rand ML, Kuhle S: Platelets and platelet function testing in children. *Prog Pediatr Cardiol* 2005;21:63–69.
- 27 Kehrel BE, Glauner M, Roberts S, Trenkmann K, Clemetson KJ, Nowak-Gottl U: Platelet functions in children from different ages. *Blood* 2001;98:251.
- 28 Gunay-Aygun M, Falik-Zaccai TC, Vilboux T, Zivony-Elboun Y, Gumruk F, Cetin M, Khayat M, Boerkoel CF, Kfir N, Huang Y, Maynard D, Dorward H, Berger K, Kleta R, Anikster Y, Arat M, Freiberg AS, Kehrel BE, Jurk K, Cruz P, Mullikin JC, White JG, Huizing M, Gahl WA: NBEAL2 is mutated in gray platelet syndrome and is required for biogenesis of platelet α -granules. *Nat Genet* 2011;43:732–734.
- 29 Jones ML, Murden SL, Bem D, Mundell SJ, Gissen P, Daly ME, Watson SP, Mumford AD; UK GAPP study group: Rapid genetic diagnosis of heritable platelet function disorders with next-generation sequencing: proof-of-principle with Hermansky-Pudlak syndrome. *J Thromb Haemost* 2012;10:306–309.
- 30 Burkhart M, Vaudel M, Gambaryan S, Radau S, Walter U, Martens L, Geiger J, Sickmann A, Zahedi RP: The first comprehensive and quantitative analysis of human platelet protein composition allows the comparative analysis of structural and functional pathways. *Blood* 2012;e73–e82.
- 31 Braunwald E, Angiolillo D, Bates E, Berger PB, Bhatt D, Cannon CP, Furman MI, Gurbel P, Michelson AD, Peterson E, Wiviott S: Assessing the current role of platelet function testing. *Clin Cardiol* 2008;31(suppl 1):110–116.
- 32 Tantry US, Gurbel PA: Assessment of oral anti-thrombotic therapy by platelet function testing. *Nat Rev Cardiol* 2011;8:572–579.
- 33 Gurbel PA, Bliden KP, Guyer K, Cho PW, Zaman KA, Kreutz RP, Bassi AK, Tantry US: Platelet reactivity in patients and recurrent events post-stenting: results of the PREPARE POST-STENTING study. *J Am Coll Cardiol* 2005;46:1820–1826.
- 34 Geisler T, Langer H, Wydymus M, Göhring K, Zürn C, Bigalke B, Stellos K, May AE, Gawaz M: Low response to clopidogrel is associated with cardiovascular outcome after coronary stent implantation. *Eur Heart J* 2006;27:2420–2425.
- 35 Geisler T, Grass D, Bigalke B, Stellos K, Drosch T, Dietz K, Herdeg C, Gawaz M: The Residual Platelet Aggregation after Deployment of Intracoronary Stent (PREDICT) score. *J Thromb Haemost* 2008;6:54–61.
- 36 Sharma RK, Voelker DJ, Sharma R, Reddy HK, Dod H, Marsh JD: Evolving role of platelet function testing in coronary artery interventions. *Vasc Health Risk Manag* 2012;8:65–75.
- 37 Collet JP, Cuisset T, Rangé G, Cayla G, Elhadad S, Pouillot C, Henry P, Motreff P, Carrié D, Boueri Z, Belle L, Van Belle E, Rousseau H, Aubry P, Monségu J, Sabouret P, O'Connor SA, Abtan J, Kerneis M, Saint-Etienne C, Barthélémy O, Beygui F, Silvain J, Vicaut E, Montalescot G; ARCTIC Investigators: bedside monitoring to adjust antiplatelet therapy for coronary stenting. *N Engl J Med* 2012;367:2100–2109.
- 38 Mehta SR, Tanguay JF, Eikelboom JW, Jolly SS, Joyner CD, Granger CB, Faxon DP, Rupprecht HJ, Budaj A, Avezum A, Widimsky P, Steg PG, Bassand JP, Montalescot G, Macaya C, Di Pasquale G, Niemela K, Ajani AE, White HD, Chrolavicius S, Gao P, Fox KA, Yusuf S; CURRENT-OASIS 7 trial investigators: Double-dose versus standard-dose clopidogrel and high-dose versus low-dose aspirin in individuals undergoing percutaneous coronary intervention for acute coronary syndromes (CURRENT-OASIS 7): a randomised factorial trial. *Lancet* 2010;376:1233–1243.
- 39 Roe MT, Armstrong PW, Fox KA, White HD, Prabhakaran D, Goodman SG, Cornel JH, Bhatt DL, Clemmensen P, Martinez F, Ardissino D, Nicolau JC, Boden WE, Gurbel PA, Ruzyllo W, Dalby AJ, McGuire DK, Leiva-Pons JL, Parkhomenko A, Gottlieb S, Topacio GO, Hamm C, Pavlides G, Goudev AR, Oto A, Tseng CD, Merkely B, Gasparovic V, Corbalan R, Cinteza M, McLendon RC, Winters KJ, Brown EB, Lokhnygina Y, Aylward PE, Huber K, Hochman JS, Ohman EM; TRILOGY ACS Investigators: Prasugrel versus clopidogrel for acute coronary syndromes without revascularization. *N Engl J Med* 2012;367:1297–1309.
- 40 Valgimigli M, Campo G, de Cesare N, Meliga E, Vranckx P, Furgieri A, Angiolillo DJ, Sabatè M, Hamon M, Repetto A, Colangelo S, Brugaletta S, Parrinello G, Percoco G, Ferrari R; Tailoring Treatment With Tirofiban in Patients Showing Resistance to Aspirin and/or Resistance to Clopidogrel (3T/2R) Investigators: Intensifying platelet inhibition with tirofiban in poor responders to aspirin, clopidogrel, or both agents undergoing elective coronary intervention: results from the double-blind, prospective, randomized Tailoring Treatment with Tirofiban in Patients Showing Resistance to Aspirin and/or Resistance to Clopidogrel study. *Circulation* 2009;119:3215–3222.
- 41 Cuisset T, Frere C, Quilici J, Morange PE, Mouret JP, Bali L, Moro PJ, Lambert M, Alessi MC, Bonnet JL: Glycoprotein IIb/IIIa inhibitors improve outcome after coronary stenting in clopidogrel nonresponders: a prospective, randomized study. *JACC Cardiovasc Interv* 2008;1:649–653.
- 42 Gibbs NM: Point-of-care assessment of antiplatelet agents in the perioperative period: a review. *Anaesth Intensive Care* 2009;37:354–369.
- 43 Breet NJ, van Werkum JW, Bouman HJ, Kelder JC, Ruven HJT, Bal ET, Deneer VH, Harmsze AM, van der Heyden JAS, Rensing BJWM, Suttorp MJ, Hackeng CM, ten Berg JM: Comparison of platelet function tests in predicting clinical outcome in patients undergoing coronary stent implantation. *JAMA* 2010;303:754–762.
- 44 Breet NJ, van Werkum JW, Bouman HJ, Kelder JC, Ruven HJ, Bal ET, Deneer VH, Harmsze AM, van der Heyden JA, Rensing BJ, Suttorp MJ, Hackeng CM, ten Berg JM: Comparison of platelet function tests in predicting clinical outcome in patients undergoing coronary stent implantation. *JAMA* 2010;303:754–762.
- 45 Panzer S, Jilma P: Methods for testing platelet function for transfusion medicine. *Vox Sang* 2011;101:1–9.
- 46 Curvers J, Dielis AW, Heeremans J, van Wersch JW: Platelet function in whole-blood donors is impaired: the effects of painkillers. *Transfusion* 2007;47:67–73.
- 47 Paglieroni TG, Janatpour K, Gosselin R, Crocker V, Dwyre DM, MacKenzie MR, Holland PV, Larkin EC: Platelet function abnormalities in qualified whole-blood donors: effects of medication and recent food intake. *Vox Sang* 2004;86:48–53.
- 48 Harrison P, Segal H, Furtado C, Verjee S, Sukhu K, Murphy MF: High incidence of defective high-shear platelet function among platelet donors. *Transfusion* 2004;44:764–770.
- 49 Pavenski K, Freedman J, Semple JW: HLA alloimmunization against platelet transfusions: pathophysiology, significance, prevention and management. *Tissue Antigens* 2012;79:237–245.
- 50 Lynen R, Legler TJ: Flow-Cytometric cross-match for the determination of platelet antibodies; in Sack U, Tárnok A, Rothe G (eds): *Cellular Diagnostics. Basics, Methods and Clinical Applications of Flow Cytometry*. Basel, Karger, 2009, pp 687–692.
- 51 Bugert P, McBride S, Smith G, Dugrillon A, Klütter H, Ouwehand WH, Metcalfe P: Microarray-based genotyping for blood groups: comparison of gene array and 5'-nuclease assay techniques with human platelet antigen as a model. *Transfusion* 2005;45:654–659.
- 52 Rinder HM, Murphy M, Mitchell JG, Stocks J, Ault KA, Hillman RS: Progressive platelet activation with storage: evidence for shortened survival of activated platelets after transfusion. *Transfusion* 1991;31:409–414.

- 53 Middelburg RA, Borkent-Raven BA, Janssen MP, van de Watering LM, Wiersum-Osselton JC, Schipperus MR, Beckers EA, Briët E, van der Bom JG: Storage time of blood products and transfusion-related acute lung injury. *Transfusion* 2012;52:658–667.
- 54 Caudrillier A, Kessenbrock K, Gilliss BM, Nguyen JX, Marques MB, Monestier M, Toy P, Werb Z, Looney MR: Platelets induce neutrophil extracellular traps in transfusion-related acute lung injury. *J Clin Invest* 2012;122:2661–2671.
- 55 Saas P, Angelot F, Bardiaux L, Seilles E, Garnache-Ottou F, Perruche S: Phosphatidylserine-expressing cell by-products in transfusion: a pro-inflammatory or an anti-inflammatory effect? *Transfus Clin Biol* 2012;19:90–97.
- 56 Jy W, Ricci M, Shariatmadar S, Gomez-Marin O, Horstman LH, Ahn YS: Microparticles in stored red blood cells as potential mediators of transfusion complications. *Transfusion* 2011;51:886–893.
- 57 Middelburg RA, Roest M, Ham J, Coccoris M, Zwaginga JJ, van der Meer PF: Flow cytometric assessment of agonist-induced P-selectin expression as a measure of platelet quality in stored platelet concentrates. *Transfusion* 2012; doi: 10.1111/trf.12001.
- 58 Cardigan R, Turner C, Harrison P: Current methods of assessing platelet function: relevance to transfusion medicine. *Vox Sang* 2005;88:153–163.
- 59 Bertolini F, Agazzi A, Peccatori F, Martinelli G, Sandri MT: The absence of swirling in platelet concentrates is highly predictive of poor posttransfusion platelet count increments and increased risk of a transfusion reaction. *Transfusion* 2000;40:121–122.
- 60 Kunicki TJ, Tuccelli M, Becker GA, Aster RH: A study of variables affecting the quality of platelets stored at 'room temperature'. *Transfusion* 1975;15: 414–421.
- 61 Gyulkhandanyan AV, Mutlu A, Freedman J, Leytin V: Markers of platelet apoptosis: methodology and applications. *J Thromb Thrombolysis* 2012;33: 397–411.
- 62 Leytin V, Freedman J: Platelet apoptosis in stored platelet concentrates and other models. *Transfus Apher Sci* 2003;28:285–295.
- 63 Estcourt LJ, Heddle N, Kaufman R, McCullough J, Murphy MF, Slichter S, Wood EM, Stanworth SJ; Biomedical Excellence for Safer Transfusion (BEST) Collaborative: The challenges of measuring bleeding outcomes in clinical trials of platelet transfusions. *Transfusion* 2013; doi: 10.1111/trf.12058.
- 64 Rinder HM, Smith BR: In vitro evaluation of stored platelets; is there hope for predicting post-transfusion platelet survival and function? *Transfusion* 2003;43:2–6.
- 65 Leytin V, Allen DJ, Mody M, Rand ML, Hannach B, Garvey B, Freedman J: A rabbit model for monitoring in vivo viability of human platelet concentrates using flow cytometry. *Transfusion* 2002;42: 711–718.
- 66 Hoffmeister KM, Felbinger TW, Falet H, Denis CV, Bergmeier W, Mayadas TN, von Andrian UH, Wagner DD, Stossel TP, Hartwig JH: The clearance mechanism of chilled blood platelets. *Cell* 2003;112:87–97.
- 67 Gitz E, Koekman CA, van den Heuvel DJ, Deckmyn H, Akkerman JW, Gerritsen HC, Urbanus RT: Improved platelet survival after cold storage by prevention of glycoprotein Ib α clustering in lipid rafts. *Haematologica* 2012;97:1873–1881.
- 68 Poston R, Gu J, Manchio J, Lee A, Brown J, Gammie J, White C, Griffith BP: Platelet function tests predict bleeding and thrombotic events after off-pump coronary bypass grafting. *Eur J Cardiothorac Surg* 2005;27:584–591.
- 69 Karger R, Reuter K, Rohlf J, Nimsky C, Sure U, Kretschmer V: The Platelet Function Analyzer (PFA-100) as a screening tool in neurosurgery. *ISRN Hematol* 2012;2012:839242.
- 70 Islam N, Fulop T, Zsom L, Miller E, Mire CD, Lebrun CJ, Schmidt DW: Do platelet function analyzer-100 testing results correlate with bleeding events after percutaneous renal biopsy? *Clin Nephrol* 2010;73:229–237.
- 71 Gogarten W, Van Aken HK: Antiplatelet drugs – implications for the anaesthesiologist. *Anesthesiol Intensivmed Notfallmed Schmerzther* 2012;47:242–251.
- 72 Gurbel PA, Mahla E, Tantry US: Peri-operative platelet function testing: the potential for reducing ischaemic and bleeding risks. *Thromb Haemost* 2011;106:248–252.
- 73 Lind SE: The bleeding time does not predict surgical bleeding. *Blood* 1991;77:2547–2552.
- 74 Rogers RPC, Levin J: A critical reappraisal of the bleeding time. *Semin Thromb Hemost* 1990;16:1–20.
- 75 Carroll RC, Chavez JJ, Snider CC, Meyer DS, Muenchen RA: Correlation of perioperative platelet function and coagulation tests with bleeding after cardiopulmonary bypass surgery. *J Lab Clin Med* 2006;147:197–204.
- 76 Korte W, Gabi K, Rohner M, Gähler A, Szadkowski C, Schnider TW, Lange J, Riesen W: Preoperative fibrin monomer measurement allows risk stratification for high intraoperative blood loss in elective surgery. *Thromb Haemost* 2005;94:211–215.
- 77 Peterson P, Hayes TE, Arkin CF, Bovill EG, Fairweather RB, Rock WA Jr, Triplett DA, Brandt JT: The preoperative bleeding time test lacks clinical benefit: College of American Pathologists' and American Society of Clinical Pathologists' position article. *Arch Surg* 1998;133:134–139.
- 78 Duke WW: The relation of blood platelets to hemorrhagic disease. *JAMA* 1910;55:1185.
- 79 Ivy AC, Nelson D, Buchet C: The standardization of certain factors in the cutaneous 'venostasis' bleeding time technique. *J Lab Clin Med* 1941;26:1812.
- 80 Lehman CM, Blaylock RC, Alexander DP, Rodgers GM: Discontinuation of the bleeding time test without detectable adverse clinical impact. *Clin Chem* 2001;47:1204–1211.
- 81 Sutor AH, Grohmann A, Kaufmehl K, Wündisch T: Problems with platelet counting in thrombocytopenia. A rapid manual method to measure low platelet counts. *Semin Thromb Hemost* 2001;27:237–243.
- 82 Harrison P, Horton A, Grant D, Briggs C, MacHin S: Immunoplatelet counting: a proposed new reference procedure. *Br J Haematol* 2000;108:228–235.
- 83 De la Salle BJ, McTaggart PN, Briggs C, Harrison P, Doré CJ, Longair I, Machin SJ, Hyde K: The accuracy of platelet counting in thrombocytopenic blood samples distributed by the UK National External Quality Assessment Scheme for General Haematology. *Am J Clin Pathol* 2012;137:65–74.
- 84 Lacroix R, Robert S, Poncelet P, Dignat-George F: Overcoming limitations of microparticle measurement by flow cytometry. *Semin Thromb Hemost* 2010;36:807–818. doi: 10.1055/s-0030-1267034.
- 85 Ruggeri ZM, Mendolicchio GL: Adhesion mechanisms in platelet function. *Circ Res* 2007;100: 1673–1685.
- 86 Stevens JM: Platelet adhesion assays performed under static conditions. *Methods Mol Biol* 2004; 272:145–151.
- 87 Curwen KD, Kim HY, Vazquez M, Handin RI, Gimbrone MA Jr: In vitro method for measuring the adherence of platelets to cultured vascular cells. *J Lab Clin Med* 1982;100:425–436.
- 88 Jurk K, Clemetson KJ, de Groot PG, Brodde MF, Steiner M, Savion N, Varon D, Sixma JJ, Van Aken H, Kehrel BE: Thrombospondin-1 mediates platelet adhesion at high shear via glycoprotein Ib (GPIb): an alternative/backup mechanism to von Willebrand factor. *FASEB J* 2003;17:1490–1492.
- 89 van Zanten GH, de Graaf S, Sloatweg PJ, Heijnen HF, Connolly TM, de Groot PG, Sixma JJ: Increased platelet deposition on atherosclerotic coronary arteries. *J Clin Invest* 1994;93:615–632.
- 90 Merten M, Dong JF, Lopez JA, Thiagarajan P: Cholesterol sulfate: a new adhesive molecule for platelets. *Circulation* 2001;103:2032–2034.
- 91 Hellem AJ: The adhesiveness of human blood platelets in vitro. *Scand J Clin Lab Invest* 1960;12 (suppl)1–117.
- 92 Salzman EW: Measurement of platelet adhesiveness. *J Lab Clin Med* 1963;62:724.
- 93 Baumgartner HR: The role of blood flow in platelet adhesion, fibrin deposition and formation of mural thrombi. *Microvasc Res* 1973;5:167–179.
- 94 Turitto VT, Baumgartner HR: Effect of physical factors on platelet adherence to subendothelium. *Thromb Diath Haemorrh* 1974;60:17–24.
- 95 Turitto VT, Baumgartner HR: Platelet interaction with subendothelium in a perfusion system: physical role of red blood cells. *Microvasc Res* 1975;9:335–344.
- 96 van Breugel HF, de Groot PG, Heethaar RM, Sixma JJ: Role of plasma viscosity in platelet adhesion. *Blood* 1992;80:953–959.
- 97 de Groot PG, Sixma JJ: Perfusion chambers; in Michelson AD (ed): *Platelets*. Academic Press, San Diego, 2002, pp 347–356.
- 98 Sixma JJ, de Groot PG, van Zanten H, IJsseldijk M: A new perfusion chamber to detect platelet adhesion using a small volume of blood. *Thromb Res* 1998;92(6 suppl 2):S43–46.
- 99 Zhao XM, Wu YP, Cai HX, Wei R, Lisman T, Han JJ, Xia ZL, de Groot PG: The influence of the pulsatility of the blood flow on the extent of platelet adhesion. *Thromb Res* 2008;121:821–825.
- 100 Westein E, de Witt S, Lamers M, Cossemans JM, Heemskerk JW: Monitoring in vitro thrombus formation with novel microfluidic devices. *Platelets* 2012;23:501–509.
- 101 Roest M, Reininger A, Zwaginga JJ, King MR, Heemskerk JW; Biorheology Subcommittee of the SSC of the ISTH: Flow chamber-based assays to measure thrombus formation in vitro: requirements for standardization. *J Thromb Haemost* 2011;9:2322–2324.
- 102 Van Kruchten R, Cossemans JM, Heemskerk JW: Measurement of whole blood thrombus formation using parallel-plate flow chambers – a practical guide. *Platelets* 2012;23:229–242.
- 103 Inoue O, Suzuki-Inoue K, McCarty OJT, Moroi M, Ruggeri ZM, Kunicki TJ, Ozaki Y, Watson SP: Laminin stimulates spreading of platelets through integrin $\alpha_6\beta_1$ -dependent activation of GPVI. *Blood* 2006;107:1405–1412.
- 104 Peters AD, Flaumenhaft R: Granule exocytosis is required for platelet spreading: differential sorting of α -granules expressing VAMP-7. *Blood* 2012;120:199–206.
- 105 Savion N, Varon D: Impact—the cone and plate(let) analyzer. *Pathophysiol Haemost Thromb* 2006;35: 83–88.
- 106 Shenkman B, Budde U, Angerhaus D, Lubetsky A, Savion N, Seligsohn U, Varon D: ADAMTS-13 regulates platelet adhesion under flow. A new method for differentiation between inherited and acquired thrombotic thrombocytopenic purpura. *Thromb Haemost* 2006;96:160–166.

- 107 Franchini M: The platelet function analyzer (PFA-100): an update on its clinical use. *Clin Lab* 2005;51:367–372.
- 108 Quiroga T, Goycoolea M, Muñoz B, Morales M, Aranda E, Panes O, Pereira J, Mezzano D: Template bleeding time and PFA-100 have low sensitivity to screen patients with hereditary mucocutaneous hemorrhages: comparative study in 148 patients. *J Thromb Haemost* 2004;2:892–898.
- 109 Sladky JL, Klima J, Grooms L, Kerlin BA, O'Brien SH: The PFA-100® does not predict delta-granule platelet storage pool deficiencies. *Haemophilia* 2012;18:626–629.
- 110 Kuijpers MJ, Heemskerk JW: Intravital imaging of thrombus formation in small and large mouse arteries: experimentally induced vascular damage and plaque rupture in vivo. *Methods Mol Biol* 2012;788:3–19.
- 111 Furie B, Furie BC: Thrombus formation in vivo. *J Clin Invest* 2005;115:3355–3362.
- 112 Kirchmaier CM, Pillitteri D: Diagnosis and management of inherited platelet disorders. *Transfus Med Hemother* 2010;37:237–246.
- 113 Summerfield GP, Keenan JP, Brodie NJ, Bellingham AJ: Bioluminescent assay of adenine nucleotides: rapid analysis of ATP and ADP in red cells and platelets using the LKB luminometer. *Clin Lab Haematol* 1981;3:257–271.
- 114 de Korte D, Gouwerok CW, Fijnheer R, Pietersz RN, Roos D: Depletion of dense granule nucleotides during storage of human platelets. *Thromb Haemost* 1990;63:275–278.
- 115 Apitz-Castro R, Cruz MR, Ledezma E, Merino F, Ramirez-Duque P, Dangelmeier C, Holmsen H: The storage pool deficiency in platelets from humans with the Chédiak-Higashi syndrome: study of six patients. *Br J Haematol* 1985;59:471–483.
- 116 Warkentin TE: HITlights: a career perspective on heparin-induced thrombocytopenia. *Am J Hematol* 2012;87(suppl 1):S92–99.
- 117 Maurer-Spurej E, Dyker K, Gahl WA, Devine DV: A novel immunocytochemical assay for the detection of serotonin in platelets. *Br J Haematol* 2002;116:604–611.
- 118 Ge S, Woo E, White JG, Haynes CL: Electrochemical measurement of endogenous serotonin release from human blood platelets. *Anal Chem* 2011;83:2598–2604.
- 119 Born GVR: Aggregation of blood platelets by adenosine diphosphate and its reversal. *Nature* 1962;194:927–929.
- 120 Morinelli TA, Niewiarowski S, Daniel JL, Smith JP: AJP receptor-mediated effects of a PGH2 analogue (U 46619) on human platelets. *Heart* 1987;253:H1035–H1043.
- 121 Kehrel B, Wierwille S, Clemetson KJ, Anders O, Steiner M, Knight CG, Fardale RW, Okuma M, Barnes MJ: Glycoprotein VI is a major collagen receptor for platelet activation: it recognizes the platelet-activating quaternary structure of collagen, whereas CD36, glycoprotein IIb/IIIa, and von Willebrand factor do not. *Blood* 1998;91:491–499.
- 122 Herczenik E, Bouma B, Korporaal SJ, Strangi R, Zeng Q, Gros P, Van Eck M, Van Berkel TJ, Gebbink MF, Akkerman JW: Activation of human platelets by misfolded proteins. *Arterioscler Thromb Vasc Biol* 2007;27:1657–1665.
- 123 Horn M, Bertling A, Brodde MF, Müller A, Roth J, Van Aken H, Jurk K, Heilmann C, Peters G, Kehrel BE: Human neutrophil alpha-defensins induce formation of fibrinogen and thrombospondin-1 amyloid-like structures and activate platelets via glycoprotein IIb/IIIa. *J Thromb Haemost* 2012;10:647–661.
- 124 Bertling A, Niemann S, Hussain M, Holbrook L, Stanley RG, Brodde MF, Pohl S, Schifferdecker T, Roth J, Jurk K, Müller A, Lahav J, Peters G, Heilmann C, Gibbins JM, Kehrel BE: Staphylococcal extracellular adherence protein induces platelet activation by stimulation of thiol isomerases. *Arterioscler Thromb Vasc Biol* 2012;32:1979–1990.
- 125 Lu Q, Clemetson JM, Clemetson KJ: Snake venoms and hemostasis. *J Thromb Haemost* 2005;3:1791–1799.
- 126 Hayward CP, Pai M, Liu Y, Moffat KA, Seecharan J, Webert KE, Cook RJ, Hedde NM: Diagnostic utility of light transmission platelet aggregometry: results from a prospective study of individuals referred for bleeding disorder assessments. *J Thromb Haemost* 2009;7:676–684.
- 127 Hayward CP, Moffat KA, Pai M, Liu Y, Seecharan J, McKay H, Webert KE, Cook RJ, Hedde NM: An evaluation of methods for determining reference intervals for light transmission platelet aggregation tests on samples with normal or reduced platelet counts. *Thromb Haemost* 2008;100:134–145.
- 128 Lages B, Weiss HJ: Biphasic aggregation responses to ADP and epinephrine in some storage pool deficient platelets: relationship to the role of endogenous ADP in platelet aggregation and secretion. *Thromb Haemost* 1980;43:147–153.
- 129 Lages B, Weiss HJ: Heterogeneous defects of platelet secretion and responses to weak agonists in patients with bleeding disorders. *Br J Haematol* 1988;68:53–62.
- 130 Olson JD, Moake JL, Collins MF, Michael BS: Adhesion of human platelets to purified solid-phase von Willebrand factor: studies of normal and Bernard-Soulier platelets. *Thromb Res* 1983;32:115–122.
- 131 Eto K, Takeshita S, Ochiai M, Ozaki Y, Sato T, Isshiki T: Platelet aggregation in acute coronary syndromes: use of a new aggregometer with laser light scattering to assess platelet aggregability. *Cardiovasc Res* 1998;40:223–229.
- 132 Grainick HR, Williams SB, McKeown LP, Rick ME, Maisonneuve P, Jenneau C, Sultan Y: Von Willebrand's disease with spontaneous platelet aggregation induced by an abnormal plasma von Willebrand factor. *J Clin Invest* 1985;76:1522–1529.
- 133 Hara K, Omori K, Sumioka Y, Aso Y: Spontaneous platelet aggregation evaluated by laser light scatter in patients with type 2 diabetes: effects of short-term improved glycemic control and adiponectin. *Transl Res* 2012;159:15–24.
- 134 Menys VC, Bhatnagar D, Mackness MI, Durrington PN: Spontaneous platelet aggregation in whole blood is increased in non-insulin-dependent diabetes mellitus and in female but not male patients with primary dyslipidemia. *Atherosclerosis* 1995;112:115–122.
- 135 Breddin HK, Lippold R, Bittner M, Kirchmaier CM, Krzywanek HJ, Michaelis J: Spontaneous platelet aggregation as a predictive risk factor for vascular occlusions in healthy volunteers? Results of the HAPARG Study. *Haemostatic parameters as risk factors in healthy volunteers. Atherosclerosis* 1999;144:211–219.
- 136 Cardinal DC, Flower RJ: The electronic aggregometer: a novel device for assessing platelet behavior in blood. *J Pharmacol Methods* 1980;3:135–158.
- 137 Fritsma GA: Platelet function testing: aggregometry and lumiaggregometry. *Clin Lab Sci* 2007;20:32–37.
- 138 Mackie IJ, Jones R, Machin SJ: Platelet impedance aggregation in whole blood and its inhibition by antiplatelet drugs. *J Clin Pathol* 1984;37:874–878.
- 139 Ingberman-Wojenski C, Smith JB, Silver MJ: Evaluation of electrical aggregometry: comparison with optical aggregometry, secretion of ATP, and accumulation of radiolabeled platelets. *J Lab Clin Med* 1983;101:44–52.
- 140 Swart SS, Pearson D, Wood JK, Barnett DB: Effects of adrenaline and alpha adrenergic antagonists on platelet aggregation in whole blood: evaluation of electrical impedance aggregometry. *Thromb Res* 1984;36:411–418.
- 141 Tóth O, Calatzis A, Penz S, Losonczy H, Siess W: Multiple electrode aggregometry: a new device to measure platelet aggregation in whole blood. *Thromb Haemost* 2006;96:781–788.
- 142 Lennon MJ, Gibbs NM, Weightman WM, McGuire D, Michalopoulos N: A comparison of Plateletworks and platelet aggregometry for the assessment of aspirin-related platelet dysfunction in cardiac surgical patients. *J Cardiothorac Vasc Anesth* 2004;18:136–140.
- 143 van Werkum JW, Harmsze AM, Elsenberg EH, Bouman HJ, ten Berg JM, Hackeng CM: The use of the VerifyNow system to monitor antiplatelet therapy: a review of the current evidence. *Platelets* 2008;19:479–488.
- 144 Smock KJ, Saunders PJ, Rodgers GM, Johari V: Laboratory evaluation of clopidogrel responsiveness by platelet function and genetic methods. *Am J Hematol* 2011;86:1032–1034.
- 145 Tschöpe D, Rösen P, Schwippert B, Kehrel B, Schausel S, Esser J, Gries FA: Platelet analysis using flowcytometric procedures. *Platelets* 1990;1:127–133.
- 146 Jurk K, Ritter MA, Schriek C, Van Aken H, Droste DW, Ringelstein EB, Kehrel BE: Activated monocytes capture platelets for heterotypic association in patients with severe carotid artery stenosis. *Thromb Haemost* 2010;103:1193–1202.
- 147 Michelson AD: Flow cytometry: a clinical test of platelet function. *Blood* 1996;87:4925–4936.
- 148 Linden MD, Frelinger AL 3rd, Barnard MR, Przyklenk K, Furman MI, Michelson AD: Application of flow cytometry to platelet disorders. *Semin Thromb Hemost* 2004;30:501–511.
- 149 Knight CJ, Panesar M, Wright C, Clarke D, Butowski PS, Patel D, Patrinely A, Fox K, Goodall AH: Altered platelet function detected by flow cytometry effects of coronary artery disease and age. *Arterioscler Thromb Vasc Biol* 1997;17:2044–2053.
- 150 Brodde MF, Jurk K, Kehrel BE: Durchflusszytometrische Untersuchungen der Thrombozytenfunktion. *Becton Dickinson*, 2012.
- 151 Furman MI, Krueger LA, Frelinger AL 3rd, Barnard MR, Mascelli MA, Nakada MT, Michelson AD: GPIIb-IIIa antagonist-induced reduction in platelet surface factor V/Va binding and phosphatidylserine expression in whole blood. *Thromb Haemost* 2000;84:492–498.
- 152 Dörmann D, Kardoeus J, Zimmermann RE, Kehrel B: Flow cytometric analysis of agonist-induced annexin V, factor Va and factor Xa binding to human platelets. *Platelets* 1998;9:171–177.
- 153 Dörmann D, Clemetson KJ, Kehrel BE: The GPIb thrombin-binding site is essential for thrombin-induced platelet procoagulant activity. *Blood* 2000;96:2469–2478.

- 154 Jurk K, Lahav J, Van Aken H, Brodde MF, Nofer JR, Kehrel BE: Extracellular protein disulfide isomerase regulates feedback activation of platelet thrombin generation via modulation of coagulation factor binding. *J Thromb Haemost* 2011; 9:2278–2290.
- 155 Tomer A: Autoimmune thrombocytopenia: determination of platelet-specific autoantibodies by flow cytometry. *Pediatr Blood Cancer* 2006;47 (5 suppl):697–700.
- 156 Kienast J, Schmitz G: Flow cytometric analysis of thiazole orange uptake by platelets: a diagnostic aid in the evaluation of thrombocytopenic disorders. *Blood* 1990;75:116–121.
- 157 Schwarz UR, Geiger J, Walter U, Eigenthaler M: Flow cytometry analysis of intracellular VASP phosphorylation for the assessment of activating and inhibitory signal transduction pathways in human platelets 11 – definition and detection of ticlopidine/clopidogrel effects. *Thromb Haemost* 1999;82:1145–52.
- 158 Robert S, Lacroix R, Poncelet P, Harhour K, Bouriche T, Judicone C, Wischhusen J, Arnaud L, Dignat-George F: High-sensitivity flow cytometry provides access to standardized measurement of small-size microparticles – brief report. *Arterioscler Thromb Vasc Biol* 2012;32:1054–1058.
- 159 Qiao JL, Shen Y, Gardiner EE, Andrews RK: Proteolysis of platelet receptors in humans and other species. *Biol Chem* 2010;391:893–900.
- 160 Duerschmied D, Canault M, Lievens D, Brill A, Cifuni SM, Bader M, Wagner DD: Serotonin stimulates platelet receptor shedding by tumor necrosis factor- α -converting enzyme (ADAM17). *J Thromb Haemost* 2009;7:1163–1171.
- 161 Clemetson KJ, Naim HY, Lüscher EF: Relationship between glyocalicin and glycoprotein Ib of human platelets. *Proc Natl Acad Sci U S A* 1981; 78:2712–2716.
- 162 Beer JH, Büchi L, Steiner B: Glyocalicin: a new assay – the normal plasma levels and its potential usefulness in selected diseases. *Blood* 1994;83: 691–702.
- 163 Forestier M, Al-Tamimi M, Gardiner EE, Hermann C, Meyer SC, Beer JH: Diesel exhaust particles impair platelet response to collagen and are associated with GPIIb/IIIa shedding. *Toxicol In Vitro*. 2012;26:930–938.
- 164 Handberg A, Norberg M, Stenlund H, Hallmans G, Attermann J, Eriksson JW: Soluble CD36 (sCD36) clusters with markers of insulin resistance, and high sCD36 is associated with increased type 2 diabetes risk. *J Clin Endocrinol Metab* 2010;95:1939–1946.
- 165 Aleil B, Kessler L, Meyer N, Wiesel ML, Simeoni J, Cazenave JP, Pinget M, Gachet C, Lanza F: Plasma levels of soluble platelet glycoprotein V are linked to fasting blood glucose in patients with type 2 diabetes. *Thromb Haemost* 2008;100:713–715.
- 166 Atalar E, Haznedaroglu IC, Kilic H, Ozer N, Coskun S, Ozturk E, Aksoyok S, Ovunc K, Kirazli S, Ozmen F: Increased soluble glycoprotein V concentration during the acute onset of unstable angina pectoris in association with chronic cigarette smoking. *Platelets* 2005;16:329–333.
- 167 Wolff V, Aleil B, Giroud M, Lorenzini JL, Meyer N, Wiesel ML, Cazenave JP, Lanza F: Soluble platelet glycoprotein V is a marker of thrombosis in patients with ischemic stroke. *Stroke* 2005;36: e17–19.
- 168 Bigalke B, Langer H, Geisler T, Lindemann S, Gawaz M: Platelet glycoprotein VI: a novel marker for acute coronary syndrome. *Semin Thromb Hemost* 2007;33:179–184.
- 169 Sahler J, Spinelli S, Phipps R, Blumberg N: CD40 ligand (CD154) involvement in platelet transfusion reactions. *Transfus Clin Biol* 2012;19:98–103.
- 170 Nesbitt WS, Harper IS, Schoenwaelder SM, Yuan Y, Jackson SP: A live cell micro-imaging technique to examine platelet calcium signaling dynamics under blood flow. *Methods Mol Biol* 2012; 788:73–89. d
- 171 Ohlmann P, Hechler B, Cazenave JP, Gachet C: Measurement and manipulation of $[Ca^{2+}]_i$ in suspensions of platelets and cell cultures. *Methods Mol Biol* 2004;273:229–250.
- 172 Mason MJ, Mahaut-Smith MP: Measurement and manipulation of intracellular Ca^{2+} in single platelets and megakaryocytes. *Methods Mol Biol* 2004; 273:251–276.
- 173 Woulfe D, Yang J, Prevost N, O'Brien P, Fortna R, Tognolini M, Jiang H, Wu J, Brass LF: Signaling receptors on platelets and megakaryocytes. *Methods Mol Biol* 2004;273:3–32.
- 174 Gibbins JM: Study of tyrosine kinases and protein tyrosine phosphorylation; in Gibbins JM, Mahut-Smith MP (eds): *Platelets and Megakaryocytes*, Volume 2: Perspectives and Techniques Series. Heidelberg, Springer, *Methods Mol Biol*, 2004; vol 273, pp 153–167.
- 175 Jones ML, Poole AW: Protein Tyrosine phosphatases; Gibbins JM, Mahut-Smith MP (eds): *Platelets and Megakaryocytes*, Volume 2: Perspectives and Techniques Series. Heidelberg, Springer, *Methods Mol Biol*, 2004; vol 273, pp 169–177.
- 176 Gorter G, Akkerman JW: The Study of serine-threonine kinases; Gibbins JM, Mahut-Smith MP (eds): *Platelets and Megakaryocytes*, Volume 2: Perspectives and Techniques Series. Heidelberg, Springer, *Methods Mol Biol*, 2004; vol 273, pp 179–199.
- 177 Shrimpton CN, Goussset K, Tablin F, Lopez JA: Isolation and analysis of platelet lipid rafts; Gibbins JM, Mahut-Smith MP (eds): *Platelets and Megakaryocytes*, Volume 2: Perspectives and Techniques Series. Heidelberg, Springer, *Methods Mol Biol*, 2004; vol 273, pp 213–228.
- 178 Naseem KM, Riba R, Troxler M: Evaluation of nitrotyrosine-containing proteins in blood platelets; Gibbins JM, Mahut-Smith MP (eds): *Platelets and Megakaryocytes*, Volume 2: Perspectives and Techniques Series. Heidelberg, Springer, *Methods Mol Biol*, 2004; vol 273, pp 301–312.
- 179 Pyne NJ, Pyne S: Receptor tyrosine kinase-G-protein-coupled receptor signalling platforms: out of the shadow? *Trends Pharmacol Sci* 2011;32: 443–50.
- 180 Chatterjee MS, Purvis JE, Brass LF, Diamond SL: Pairwise agonist scanning predicts cellular signaling responses to combinatorial stimuli. *Nat Biotechnol* 2010;28:727–732.
- 181 Boyanova D, Nilla S, Birschmann I, Dandekar T, Dittrich M: PlateletWeb: a systems biologic analysis of signaling networks in human platelets. *Blood* 2012;119:e22.
- 182 Dittrich M, Birschmann I, Mietner S, Sickmann A, Walter U, Dandekar T: Platelet protein interactions: map, signaling components and phosphorylation groundstate. *Arterioscler Thromb Vasc Biol* 2008;28:1326–1331.
- 183 Schulz C, Leuschen NV, Fröhlich T, Lorenz M, Pfeiler S, Gleissner CA, Kremmer E, Kessler M, Khandoga AG, Engelmann B, Ley K, Massberg S, Arnold GJ: Identification of novel downstream targets of platelet glycoprotein VI activation by differential proteome analysis: implications for thrombus formation. *Blood* 2010;115:4102–4110.
- 184 Zellner M, Baureder M, Rappold E, Bugert P, Kotzailias N, Babeluk R, Baumgartner R, Attems J, Gerner C, Jellinger K, Roth E, Oehler R, Umlauf E: Comparative platelet proteome analysis reveals an increase of monoamine oxidase-B protein expression in Alzheimer's disease but not in non-demented Parkinson's disease patients. *J Proteomics* 2012;75:2080–2092.
- 185 Della Corte A, Tamburrelli C, Crescente M, Giordano L, D'Imperio M, Di Michele M, Donati MB, De Gaetano G, Rotilio D, Cerletti C: Platelet proteome in healthy volunteers who smoke. *Platelets* 2012;23:91–105.
- 186 Della Corte A, Maugeri N, Pampuch A, Cerletti C, de Gaetano G, Rotilio D: Application of 2-dimensional difference gel electrophoresis (2D-DIGE) to the study of thrombin-activated human platelet secretome. *Platelets* 2008;19:43–50.
- 187 Senis YA, Tomlinson MG, García A, Dumon S, Heath VL, Herbert J, Cobbold SP, Spalton JC, Ayman S, Antrobus R, Zitzmann N, Bicknell R, Frampton J, Authi KS, Martin A, Wakelam MJ, Watson SP: A comprehensive proteomics and genomics analysis reveals novel transmembrane proteins in human platelets and mouse megakaryocytes including G6b-B. *Mol Cell Proteomics* 2007;6:548–564.
- 188 Schweigel H, Geiger J, Beck F, Buhs S, Gerull H, Walter U, Sickmann A, Nollau P: Deciphering of ADP-induced, phosphotyrosine-dependent signaling networks in human platelets by SH2 profiling. *Proteomics* 2013; doi: 10.1002/pmic.201200353.
- 189 García A: Proteome analysis of signaling cascades in human platelets. *Blood Cells Mol Dis* 2006;36: 152–156.
- 190 Dowal L, Yang W, Freeman MR, Steen H, Flaumenhaft R: Proteomics analysis of palmitoylated platelet proteins. *Blood* 2011;118:e62–e73.
- 191 Parguina AF, Rosa I, García A: Proteomics applied to the study of platelet-related diseases: aiding the discovery of novel platelet biomarkers and drug targets. *J Proteomics* 2012;76:275–286.
- 192 Parguina AF, Grigorian-Shamajian L, Agra RM, Teixeira-Fernández E, Rosa I, Alonso J, Viñuela-Roldán JE, Seoane A, González-Juanatey JR, García A: Proteins involved in platelet signaling are differentially regulated in acute coronary syndrome: a proteomic study. *PLoS One* 2010;5(10): e13404.
- 193 Randriamboavonjy V, Isaak J, Elghezawy A, Pistrosch F, Frömel T, Yin X, Badenhop K, Heide H, Mayr M, Fleming I: Calpain inhibition stabilizes the platelet proteome and reactivity in diabetes. *Blood* 2012;120:415–423.
- 194 Di Michele M, Thys C, Waelkens E, Overbergh L, D'Hertog W, Mathieu C, De Vos R, Peerlinck K, Van Geet C, Freson K: An integrated proteomics and genomics analysis to unravel a heterogeneous platelet secretion defect. *J Proteomics* 2011;74:902–913.
- 195 Burkhart JM, Vaudel M, Gambaryan S, Radau S, Walter U, Martens L, Geiger J, Sickmann A, Zahedi RP: The first comprehensive and quantitative analysis of human platelet protein composition allows the comparative analysis of structural and functional pathways. *Blood* 2012;120:e73–82.
- 196 Nurden AT, Fiore M, Pillois X, Nurden P: Genetic testing in the diagnostic evaluation of inherited platelet disorders. *Semin Thromb Hemost* 2009;35:204–212.
- 197 Leytin V: Apoptosis in the anucleate platelet. *Blood Rev* 2012;26:51–63.

- 198 Winkler J, Rand ML, Schmutge M, Speer O: Omi/HtrA2 and XIAP are components of platelet apoptosis signalling. *Thromb Haemost* 2013;109:532–539.
- 199 Leytin V, Freedman J: Platelet apoptosis in stored platelet concentrates and other models. *Transfus Apher Sci* 2003;28:285–295.
- 200 C. Bertino AM, Qi XQ, Li J, Xia Y, Kuter DJ: Apoptotic markers are increased in platelets stored at 37 degrees. *Transfusion* 2003;43:857–866.
- 201 Perrotta PL, Perrotta CL, Snyder EL: Apoptotic activity in stored human platelets. *Transfusion* 2003;43:526–535.
- 202 Leytin V, Allen DJ, Lyubimov E, Freedman J: Higher thrombin concentrations are required to induce platelet apoptosis than to induce platelet activation. *Br J Haematol* 2007;136:762–764.
- 203 Leytin V, Allen DJ, Mutlu A, Gyulhandyan AV, Mykhaylov S, Freedman J: Mitochondrial control of platelet apoptosis: effect of cyclosporin A, an inhibitor of the mitochondrial permeability transition pore. *Lab Invest* 2009;89:374–384.
- 204 Monroe DM, Hoffman M, Roberts HR: Platelets and thrombin generation. *Arterioscler Thromb Vasc Biol* 2002;22:1381–1389.
- 205 Jackson SP: Arterial thrombosis – insidious, unpredictable and deadly. *Nat Med* 2011;17:1423–1436.
- 206 Schoenwaelder SM, Yuan Y, Josefsson EC, White MJ, Yao Y, Mason KD, O'Reilly LA, Henley KJ, Ono A, Hsiao S, Willcox A, Roberts AW, Huang DC, Salem HH, Kile BT, Jackson SP: Two distinct pathways regulate platelet phosphatidylserine exposure and procoagulant function. *Blood* 2009;114:663–666.
- 207 Whelihan MF, Mann KG: The role of the red cell membrane in thrombin generation. *Thromb Res* 2013; doi: 10.1016/j.thromres.2013.01.023.
- 208 Heemskerk JW, Bevers EM, Lindhout T: Platelet activation and blood coagulation. *Thromb Haemost* 2002;88:186–193.
- 209 Jurk K, Kehrel BE: Platelets: physiology and biochemistry. *Semin Thromb Hemost* 2005;31:381–392.
- 210 Cosemans JM, Schols SE, Stefanini L, de Witt S, Feijge MA, Hamulyak K, Deckmyn H, Bergmeier W, Heemskerk JW: Key role of glycoprotein Ib/V/IX and von Willebrand factor in platelet activation-dependent fibrin formation at low shear flow. *Blood* 2011;117: 651–660.
- 211 Tait JF, Smith C, Wood BL: Measurement of phosphatidylserine exposure in leukocytes and platelets by whole-blood flow cytometry with annexin V. *Blood Cells Mol Dis* 1999;25:271–278.
- 212 Stavenuiter F, Davis NF, Duan E, Gale AJ, Heeb MJ: Platelet protein S directly inhibits procoagulant activity on platelets and microparticles. *Thromb Haemost* 2013;109:229–237.
- 213 Duckers C, Simioni P, Spiezia L, Radu C, Dabrilili P, Gavasso S, Rosing J, Castoldi E: Residual platelet factor V ensures thrombin generation in patients with severe congenital factor V deficiency and mild bleeding symptoms. *Blood* 2010; 115:879–886.
- 214 Dale GL, Friese P, Batar P, Hamilton SF, Reed GL, Jackson KW, Clemetson KJ, Alberio L: Stimulated platelets use serotonin to enhance their retention of procoagulant proteins on the cell surface. *Nature* 2002;415:175–179.
- 215 Prodan CI, Szasz R, Vincent AS, Ross ED, Dale GL: Coated-platelets retain amyloid precursor protein on their surface. *Platelets* 2006;17:56–60.
- 216 Prodan CI, Ross ED, Vincent AS, Dale GL: Rate of progression in Alzheimer's disease correlates with coated-platelet levels – a longitudinal study. *Transl Res* 2008;152:99–102.
- 217 Prodan CI, Ross ED, Stoner JA, Cowan LD, Vincent AS, Dale GL: Coated-platelet levels and progression from mild cognitive impairment to Alzheimer disease. *Neurology* 2011;76:247–252.
- 218 Prodan CI, Ross ED, Vincent AS, Dale GL: Coated-platelets are higher in amnesic versus non-amnesic patients with mild cognitive impairment. *Alzheimer Dis Assoc Disord* 2007;21:259–261.
- 219 Prodan CI, Vincent AS, Dale GL: Coated platelet levels correlate with bleed volume in patients with spontaneous intracerebral hemorrhage. *Stroke* 2010;41:1301–1303.
- 220 Prodan CI, Joseph PM, Vincent AS, Dale GL: Coated-platelets in ischemic stroke: differences between lacunar and cortical stroke. *J Thromb Haemost* 2008;6:609–614.
- 221 Hemker HC, Giesen PL, Ramjee M, Wagenvoort R, Béguin S: The thrombogram: monitoring thrombin generation in platelet rich plasma. *Thromb Haemost* 2000;83:589–591.
- 222 Hemker HC, Giesen P, Aldieri R, Regnault V, de Smed E, Wagenvoort R, Lecompte T, Béguin S: The calibrated automated thrombogram (CAT): a universal routine test for hyper- and hypocoagulability. *Pathophysiol Haemost Thromb* 2002;2: 249–253.
- 223 Semeraro F, Ammolio CT, Morrissey JH, Dale GL, Friese P, Esmon NL, Esmon CT: Extracellular histones promote thrombin generation through platelet-dependent mechanisms: involvement of platelet TLR2 and TLR4. *Blood* 2011; 118:1952–1961.
- 224 Schwertz H, Rowley JW, Tolley ND, Campbell RA, Weyrich AS: Assessing protein synthesis by platelets; in Gibbins JM, Mahut-Smith MP (eds): *Platelets and Megakaryocytes*, Volume 3, Additional Protocols and Perspectives. Heidelberg, Springer, Methods Mol Biol. 2012, 788, pp 141–153.
- 225 Feghhi S, Sniadecki NJ: Chanobiology of platelets: techniques to study the role of fluid flow and platelet retraction forces at the micro- and nanoscale. *Int J Mol Sci* 2011;12:9009–9030.
- 226 Carr ME: Development of platelet contractile force as a research and clinical measure of platelet function. *Cell Biochem Biophys* 2003;38:55–78.
- 227 Ono A, Westein E, Hsiao S, Nesbitt WS, Hamilton JR, Schoenwaelder SM, Jackson SP: Identification of a fibrin-independent platelet contractile mechanism regulating primary hemostasis and thrombus growth. *Blood* 2008;112:90–99.
- 228 Lam WA, Chaudhuri O, Crow A, Webster KD, Li TD, Kita A, Huang J, Fletcher DA: Mechanics and contraction dynamics of single platelets and implications for clot stiffening. *Nat Mater* 2011; 10:61–66.
- 229 Schwarz Henriques S, Sandmann R, Strate A, Köster S: Force field evolution during human blood platelet activation. *J Cell Sci* 2012;125: 3914–3920.
- 230 Carlsson L, Markey F, Blikstad I, Persson T, Lindberg U: Reorganization of actin in platelets stimulated by thrombin as measured by the DNase I inhibition assay. *Proc Natl Acad Sci U S A* 1979;76:6376–6380.
- 231 Cassimeris L, McNeill H, Zigmond SH: Chemoattractant-stimulated polymorphonuclear leukocytes contain two populations of actin filaments that differ in their spatial distributions and relative stabilities. *J Cell Biol* 1990;110:1067–1075.
- 232 Nachmias VT, Cassimeris L, Golla R, Safer D: Thymosin beta 4 (T beta 4) in activated platelets. *Eur J Cell Biol* 1993;61:314–320.
- 233 McCarty OJ, Larson MK, Auger JM, Kalia N, Atkinson BT, Pearce AC, Ruf S, Henderson RB, Tybulewicz VL, Machesky LM, Watson SP: Rac1 is essential for platelet lamellipodia formation and aggregate stability under flow. *J Biol Chem* 2005;280:39474–39484.
- 234 Leng L, Kashiwagi H, Ren XD, Shattil SJ: RhoA and the function of platelet integrin alphaIIb-beta3. *Blood* 1998;91:4206–4215.
- 235 Shah MD, Bergeron AL, Dong JF, López JA: Flow cytometric measurement of microparticles: pitfalls and protocol modifications. *Platelets* 2008; 19:365–372.
- 236 Khorana AA, Francis CW, Menzies KE, Wang JG, Hyrien O, Hathcock J, Mackman N, Taubman MB: Plasma tissue factor may be predictive of venous thromboembolism in pancreatic cancer. *J Thromb Haemost* 2008;6:1983–1985.
- 237 Lacroix R, Judicone C, Poncelet P, Robert S, Arnaud L, Sampol J, Dignat-George F: Impact of pre-analytical parameters on the measurement of circulating microparticles: towards standardization of protocol. *J Thromb Haemost* 2012;10:437–446.
- 238 Trummer A, De Rop C, Tiede A, Ganser A, Eisert R: Recovery and composition of microparticles after snap-freezing depends on thawing temperature. *Blood Coagul Fibrinolysis*. 2009;20:52–56.
- 239 White JG, Gerrard JM: The ultrastructure of defective human platelets. *Mol Cell Biochem* 1978; 21:109–128.
- 240 Nofer JR, Herminghaus G, Brodde M, Morgenstern E, Rust S, Engel T, Seedorf U, Assmann G, Bluethmann H, Kehrel BE: Impaired platelet activation in familial high density lipoprotein deficiency (Tangier disease). *J Biol Chem* 2004;279: 34032–34037.
- 241 Morgenstern E, Patscheke H, Mathieu G: The origin of the membrane convolute in degranulating platelets. A comparative study of normal and 'gray' platelets. *Blut* 1990;60:15–22.
- 242 Clauser S, Cramer-Bordé E: Role of platelet electron microscopy in the diagnosis of platelet disorders. *Semin Thromb Hemost* 2009;35:213–223.
- 243 Karagkiozaki V, Logothetidis S, Kalfagiannis N, Lousinian S, Giannoglou G: Atomic force microscopy probing platelet activation behavior on titanium nitride nanocoatings for biomedical applications. *Nanomedicine* 2009;5:64–72.
- 244 Leong HS, Podor TJ, Manocha B, Lewis JD: Validation of flow cytometric detection of platelet microparticles and liposomes by atomic force microscopy. *J Thromb Haemost* 2011;9:2466–2476.
- 245 Hussain MA, Siedlecki CA: The platelet integrin alpha(IIb) beta(3) imaged by atomic force microscopy on model surfaces. *Micron* 2004;35:565–573.
- 246 Baker-Groberg SM, Phillips KG, McCarty OJ: Quantification of volume, mass, and density of thrombus formation using brightfield and differential interference contrast microscopy. *J Biomed Opt* 2013;18:16014.