Research Article

New Approach to the Analysis of Raw Data from the Ocular Response Analyzer

Agnieszka Jóźwik^a Henryk Kasprzak^a Marta Kuczma^b^aDepartment of Optics and Photonics, Wrocław University of Technology, Wrocław, and^bMedical Clinic Oko-Laser, Krakow, Poland

What Is It about?

We present a pilot study describing a new method of processing and analyzing raw data from the Ocular Response Analyzer (ORA) that can be used to describe and examine some new potential physical features of the anterior eye. Multiple measurements were made on 29 patients, distinguishing between the normal and the glaucoma group. The proposed new characteristics make it possible to differentiate the healthy reference group from the glaucoma patients.

Keywords

Eye · Anterior segment · Air puff tonometry · Ocular Response Analyzer · Eye biomechanics · Corneal applanation · Raw data

Abstract

Purpose: The aim of this research was to present a new approach for the potential characteristics of the anterior eye, based on numerical analysis of the raw data from multiple measurements on 1 eye captured with use of the Ocular Response Analyzer (ORA). **Methods:** The measurements were carried out 6 times on every eye, one by one, on 29 patients using the ORA. Dependencies between values of 6 applanation pressures P_1 and P_2 from 1 eye and the new proposed time-dependent parameters sm and smm , obtained from the raw data, were used to determine new characteristic features of the anterior eye. **Results:** Obtained values for multiple applanation pressures P_1 and P_2 and their average value P_a from all measurements on 1 eye show strong linear dependencies with the new proposed time-dependent parameter smm . **Conclusions:** The new parameters based on slope coefficients a_1 and a_2 , their difference Δa , and respective y-intercept values b_1 , b_2 , and Δb show some characteristic stability for a multiply measured eye and can be used for its characterization. These parameters can be used for differentiation of the eyes with higher accuracy than those obtained from a single measurement. They are likely related to some biomechanical properties of the anterior eye.

© 2016 The Author(s)
Published by S. Karger AG, Basel

Agnieszka Jóźwik, PhD
Department of Optics and Photonics
Wrocław University of Technology
Wyb. Wyspiańskiego 27, PL-50-370 Wrocław (Poland)
E-Mail agnieszka.jozwik@pwr.edu.pl

Introduction

Intraocular pressure (IOP) is one of the parameters that are used in the diagnosis of glaucoma, in the monitoring of glaucoma progression, and in the evaluation of treatment effectiveness. The Ocular Response Analyzer (ORA; Reichert Technologies, Depew, NY, USA) is the first commercially available noncontact tonometer giving the possibility to measure, apart from Goldmann-correlated IOP (IOPg) and corneal compensated IOP (IOPcc), some specific biomechanical properties of the cornea, such as corneal hysteresis (CH) and the corneal resistance factor (CRF) [1, 2].

A single-pulse air stream from the pump bends the cornea. At the same time the corneal apex is illuminated by a laser diode, located out of axis of the measured eye. The pressure of the air puff is measured dynamically in the air jet, and the light intensity reflected from the corneal apex is synchronously detected by a photodiode, placed out of axis of the eye, symmetrically to the laser diode. The signal of the air puff pressure presents the pressure curve and consists of 400 measured points within 25 ms. According to information given by the manufacturer, the air pump is switched off after the first corneal applanation; however, the air pressure still increases for another 3–4 ms. After having reached its maximum, the pressure starts decreasing to its initial value. The shape of the air pressure curve is basically similar to a Gaussian distribution, but more detailed analysis shows that the pressure curve is not symmetrical [3]. The second curve (applanation) given by the ORA also consists of 400 points measured within 25 ms, but they represent the intensity of light reflected from the corneal center and registered by the photodiode. The applanation curve has 2 specific maximal values, where the corneal surface reaches 2 applanation states – during the inward and the outward corneal deformations. However, in many cases both applanation areas consist of a few local maxima, which usually complicates determination of the “best” applanation points. The ORA’s software smoothens the applanation curve in order to get 2 clearer and explicitly distinct maxima, which define 2 applanation times t_{1a} and t_{2a} .

The applanation pressures P_1 and P_2 are estimated for the applanation times t_{1a} and t_{2a} , from the air pressure curve [2]. The next 4 principal ORA parameters IOPg, IOPcc, CH, and CRF are calculated from these values of applanation pressures P_1 and P_2 . The ORA also gives an additional 37 parameters derived from the nonsmoothed applanation curve, which can be used for characterization of the applanation signal [4, 5]. Some authors introduce new parameters or modify the parameters given by the ORA in order to make them more useful in ophthalmic diagnosis [6–8].

The waveform score (WS) has values ranging from 0 to 10 and describes the quality of the measurement. Various lower limits of acceptable values for WS were reported, giving reliable results of measurements. Usually these limits are suggested to be within the range of 4–7 [9–11]. However, according to our personal experience and results, the WS parameter describes not only the quality of measurements, but also the quality of corneal applanation and different corneal abnormalities. Low values of WS obtained for the same patient seem to give some valuable information about the corneal condition.

Repeated measurements of IOPg, IOPcc, CH, and CRF for 1 patient give quite a significant distribution of results [12–14]. As proposed in the paper, more detailed analysis of these IOPg distributions and their dependencies on some parameters related to the geometry of both applanation and air pressure curves can manifest some interesting correlations. These dependencies can be used in determination of new parameters describing the properties of the cornea and the anterior eye.

This work presents a pilot study on new dependencies between parameters estimated from the raw data of both curves given by the ORA, and their potential application in determination of biomechanical properties of the anterior eye.

Fig. 1. Data points for the region of the first applanation.

Materials and Methods

ORA measurements were repeatedly performed on 29 volunteers (aged 25–78 years, mean $\pm$ SD 51.4 ± 18.5 years) between January and March 2013. Ten patients (aged 25–39 years, mean $\pm$ SD 32.2 ± 5.0 years) without identification of glaucoma were classified as the normal group, while 19 subjects (aged 32–78 years, mean $\pm$ SD 61.6 ± 14.3 years) were classified as the glaucomatous group. The ORA gives 4 results of the measured parameter, which are usually averaged by the examiner. In the current study a different approach is proposed: more measurements were performed on each patient, and some characteristic features of these measurements were taken for further consideration and analysis. In order not to cause discomfort to the subjects, 6 measurements were performed on each examined eye. However, in order to test the repeatability of examined features, the eyes of 9 subjects from the healthy group were measured 15 times. All measurements were performed within 1 day, with at least 15-min intervals between the consecutive measurements. All patients were fully informed about the details of these noninvasive procedures as well as the purpose and possible consequences of the study. Conscious agreement was obtained from the subjects. The research followed the tenets of the Declaration of Helsinki and was approved by the Ethics Committee of the Medical University of Wrocław (KB 503/2011).

The results were repeatable even for lower values of WS that are usually eliminated [9–11]. However, WS measurements below 2 were rejected due to difficulty of numerical analysis, caused by significantly increasing dispersal of calculated parameters. The raw data points of the air pressure and applanation curves, exported directly from the instrument, were numerically processed in a MATLAB environment (Math Works, Inc., Natick, MA, USA, version 2011a).

In order to find the optimal value of both the first t_{1a} and the second t_{2a} applanation times from the raw data, 5 different methods were used to estimate the timing of both applanations.

The first 2 methods were based on calculation of the center of mass (centroid) of the 2 areas of the applanation curve close to the applanation region. The applanation region was always selected as the region where the values of the applanation curve exceeded 70% of the maximal peak for the respective applanation. The data points of the first applanation region and their smoothing are shown in Figure 1. In the first method, the centroid (representing

Fig. 2. Considered parameters describing air pressure and appplanation curves.

times t_{1a} or t_{2a}) was calculated only from the local maxima in the established appplanation region, while all the data points from this region were taken into account for centroid calculation in the second method. The third and fourth methods used the same data points as the first and the second method, but instead of centroids, the centers of inertia were calculated. The fifth method estimated the appplanation time from the maximum of the smoothed appplanation curve. The original, raw appplanation curve was smoothed by means of a Gaussian filter with the window size equal to 17 points. The size of the window was selected from the highest correlation, presented in the next section. The respective values of the appplanation pressures P_1 and P_2 were calculated for each measurement from appplanation times t_{1a} and t_{2a} , obtained for all 5 methods.

While testing and estimating a number of different dependencies between both the appplanation air pressures P_1 , P_2 , or their arithmetic average P_a and time-related parameters of the air pressure and appplanation curves, some interesting correlations were observed. The highest, unexpected characteristic correlations for 1 patient were observed between pressures P_1 , P_2 , or P_a , and two time-related parameters of both ORA curves, defined as

$$sm = \frac{t_m}{dm}, \quad (1)$$

$$smm = \frac{t_m}{\sqrt{dm_1 \times dm_2}} - 1, \quad (2)$$

the time periods t_m , dm_1 , and dm_2 being shown in Figure 2. The parameter t_m is the time of maximum air pressure, and dm is the time between two appplanations (sum of dm_1 and dm_2).

The relationships between the values P_1 , P_2 , or P_a and smm or sm show strong linear dependency for each eye under examination. The following parameters were calculated for all dependencies: slope a , y-intercept b and the Pearson correlation coefficient R . The calculations were performed for all the 5 methods of appplanation time determination. Statistical analysis of linear dependency was performed using 2-tailed test significance of determination. A p value <0.05 indicates a statistically significant difference between 2 parameters.

Table 1. Values of Pearson correlation coefficients R for 29 patients for dependencies between pressures P_1 , P_2 , or P_a and sm or smm for 29 patients

Patient No.	P_1 (sm)	P_2 (sm)	P_a (sm)	P_1 (smm)	P_2 (smm)	P_a (smm)
1	0.875	0.971	0.992	0.885	0.962	0.992
2	0.796	0.986	0.969	0.823	0.978	0.977
3	0.793	0.920	0.979	0.787	0.921	0.977
4	0.998	0.989	0.997	0.998	0.987	0.996
5	0.914	0.967	0.975	0.921	0.966	0.977
6	0.775	0.993	0.968	0.890	0.986	0.992
7	0.978	0.978	0.993	0.964	0.979	0.994
8	0.974	0.986	0.994	0.986	0.965	0.989
9	0.915	0.864	0.995	0.894	0.883	0.995
10	0.895	0.979	0.968	0.906	0.975	0.972
11	0.997	0.977	0.995	0.995	0.967	0.991
12	0.977	0.927	0.974	0.984	0.922	0.975
13	0.971	0.958	0.999	0.969	0.960	0.997
14	0.888	0.750	0.997	0.899	0.654	0.989
15	0.959	0.887	0.989	0.954	0.627	0.951
16	0.877	0.540	0.988	0.908	0.608	0.970
17	0.720	0.903	0.958	0.875	0.933	0.974
18	0.574	0.757	0.921	0.425	0.815	0.932
19	0.997	0.998	0.999	0.999	0.998	0.999
20	0.865	0.951	0.988	0.957	0.888	0.962
21	0.838	0.854	0.977	0.889	0.904	0.987
22	0.907	0.972	0.994	0.880	0.936	0.995
23	0.656	0.970	0.991	0.784	0.732	0.967
24	0.942	0.940	0.974	0.768	0.892	0.916
25	0.975	0.984	0.992	0.984	0.970	0.992
26	0.843	0.946	0.970	0.851	0.953	0.977
27	0.987	0.997	0.999	0.988	0.998	0.998
28	0.975	0.937	0.981	0.963	0.947	0.984
29	0.892	0.909	0.990	0.912	0.897	0.999

Results

Linear dependencies with high correlation coefficients and p values <0.05 between pressures of applanation P_1 , P_2 , P_a and parameters sm or smm were obtained for all 29 patients with repetition of 6 measurements. Pearson correlation coefficients are presented in Table 1. The results show that the difference between calculations of parameters from 15 and 6 measurements was not statistically significant ($p > 0.05$). Numerical analysis showed that correlations between both pressures P_1 or P_2 and the smm parameter were slightly larger than the correlations between both pressures and sm for the majority of measurements. The differences in the obtained correlations calculated for the 5 methods of the estimation of applanation timing described above were small, and there was no statistical difference (Wilcoxon test) between them ($p > 0.05$). Significant differences were obtained for signals with $WS < 2$, but these data were not used for further analysis. The first method of calculation of both applanation times with the centroid of the local maxima was chosen for further consideration.

Some exemplary results obtained for 2 patients are shown in Figure 3. High correlations between P_1 , P_2 , and smm can be observed. However, the highest correlation was always between P_a and smm ($R > 0.95$). Such a very high correlation is rarely observed in measure-

Fig. 3. Dependencies between respective appplanation pressure values (P_1 , highest line; P_2 , lowest line; P_a , central line) and smm parameter for 2 exemplary patients.

ments of biological parameters. An additional interesting feature is that the dispersion of data points for P_a from the straight line is lower than those for both the P_1 and P_2 lines in all cases; the respective correlation coefficient was close to 1. It is worth noticing that such high correlations appeared even for relatively low values of WS. Slopes a_{P1} , a_{P2} , and a_a and the respective y-intercept values b_{P1} , b_{P2} , and b_a were calculated for all patients, for the 3 dependencies presented in Figure 3. Differences between the slopes and the y-intercept values of both curves representing pressures P_1 and P_2 were calculated as follows:

$$\Delta a = a_{P1} - a_{P2}, \quad (3)$$

$$\Delta b = b_{P1} - b_{P2}. \quad (4)$$

It seems to be expected that according to hysteresis properties, both the P_1 and P_2 lines should be more or less parallel. However, our analysis showed that for some patients these 2 lines were converging and for the other diverging, giving positive or negative differences Δa . The values of Δa were distributed almost symmetrically with relation to zero. The distribution of Δa values was close to a normal distribution at the level of significance $\alpha = 0.05$ (Shapiro-Wilk test). For more than 6 measurements on 1 eye, the correlations between the parameters were comparable. However, by increasing the number of repeated measurements on 1 eye, the examination becomes more uncomfortable for the patient.

Fig. 4. Dependencies between Δa and a_{p1} (dashed line, circles) as well as a_{p2} (solid line, crosses) for 29 patients.

It was noticeable that Δa was strongly dependent on the values of a_{p1} ($R = 0.913$, $p < 0.001$) and a_{p2} ($R = -0.950$, $p < 0.001$). Figure 4 presents the values of slopes a_{p1} and a_{p2} plotted as a function of the slope differences Δa , as well as a high correlation between Δa and y-intercepts b_{p1} ($R = 0.893$) and b_{p2} ($R = 0.957$). All dependencies were statistically significant. There was no significant correlation between Δa and a_a ($R = 0.375$) and b_a ($R = 0.458$). The analysis of the Δb parameter also manifested a very high correlation to the parameters: a_{p1} ($R = 0.910$), a_{p2} ($R = 0.929$), b_{p1} ($R = 0.922$), and b_{p2} ($R = 0.958$). The Shapiro-Wilk test showed that there is no reason to reject the hypothesis of a lack of normality of the distribution of Δb values. Similar to the case of the slope differences Δa , there were low Pearson correlation coefficients between Δb and a_a ($R = 0.347$) or b_a ($R = 0.426$) values. It is worth noticing that Δa and Δb were better correlated to the parameters related to the second applanation (a_{p2} and b_{p2}) than to the respective parameters of the first applanation (a_{p1} and b_{p1}).

The values of the slope a_{p1} showed lack of correlation to the slope of average pressures a_a ($R = 0.035$, $p = 0.86$), while the values of the slope a_{p2} were moderately correlated to a_a ($R = 0.646$, $p < 0.001$). Similar results were observed for y-intercept values b_{p1} versus b_a ($R = -0.043$, $p = 0.82$) and b_{p2} versus b_a ($R = 0.668$, $p < 0.001$).

Further analysis showed that the values Δa were neither correlated to the average values of CH ($R < 0.1$) nor to the values of IOPg ($R < 0.1$). It is likely that the value Δa is related to some other biomechanical features of the anterior eye.

Figure 5 presents the high correlation between Δa and b_{p1} ($R = -0.893$, $p < 0.001$). One could notice that there was a higher concentration of results for healthy subjects in the center of the graph, while the results obtained for glaucomatous patients manifested larger dispersion. However, in some cases the results for the glaucomatous patients were also localized in the central part of the graph.

Interesting results can be observed in Figure 6, which presents the values of b_a plotted against the values of a_a . In general, all results can be approximated by a linear function (continuous line) with a relatively high correlation ($R = -0.884$, $p < 0.001$). If one separates the healthy group (circles) from the glaucomatous patients (crosses), it can be observed that both groups can be approximated by 2 different lines with significantly higher correlations ($R = -0.978$ for normal subjects and $R = -0.981$ for glaucomatous patients). Linear approximation can be written as

$$b_a = -1.42 \times a_a + 176 \quad (5)$$

Fig. 5. Dependencies between Δa and b_{p1} for normal (circles) and glaucoma (crosses) patients.

Fig. 6. Variability of y-intercept b_a parameter versus average slope a_a , separated into 2 groups.

for the normal group, and

$$b_a = -1.44 \times a_a + 200 \quad (6)$$

for the glaucomatous patients. Statistical analysis showed significant differences in the parameters of these lines.

The difference between the 2 groups was particularly visible in the 3D graph, where the respective values of Δa were taken into account and plotted as the third axis (Fig. 7). The respective points representing the 2 groups of patients were located in 2 different planes.

The planar approximation on a 3D graph for the glaucoma patients is represented by the equation:

$$b_a = 0.141 \times \Delta a - 1.33 \times a_a + 168, \quad (7)$$

with multiple correlation coefficient $R = 0.996$ while for the healthy subjects:

$$b_a = -0.018 \times \Delta a - 1.50 \times a_a + 209, \quad (8)$$

with respective multiple correlation coefficient $R = 0.984$.

Fig. 7. 3D graph of b_a values versus slope difference Δa and slope a_a , with respective planar approximations for 2 groups of patient with their projections on 3 respective planes.

Fig. 8. Results of measurements and linear regression for dependency between IOPg obtained from the ORA and P_a (black line) as well as the line defined by equation 10 (gray line).

In order to examine the correlation between the average value of both applanation pressures P_a obtained from raw data and the value of IOPg given by the ORA, all measurements for 29 patients (29×6) were taken into account. Figure 8 shows a very high correlation between the examined values ($R = 0.986$, $p < 0.001$). Estimated linear approximation is given by the equation

$$IOPg = P_a \times 0.134 - 10.5. \quad (9)$$

However, the equation given by the ORA's manufacturer [1] for the same dependency differs from the one above:

$$IOPg = P_a \times 0.149 - 15. \quad (10)$$

Both lines are marked in Figure 8. One can see that the results for the healthy subjects (crosses) characterize higher dispersion in comparison to those of glaucoma patients (squares), but both of them are not located on the line defined by equation 9.

Discussion

This paper proposes a new method of processing and analysis of the raw data from the ORA instrument. According to the ORA's software suggestion, usually 1–4 measurements are taken on 1 eye, and the results for the highest values of WS parameter are the final results of the measurement. However, consecutive repetition of the measurements on the same eye can give quite a wide distribution of obtained results. Interpretation of this distribution is still not clear. It is very likely that the anterior eye changes its physical features during quasi-periodical IOP variations due to blood pulsation [15]. Usually, the values of IOPg given by the ORA are averaged, and its mean value is treated as the closest to the real one. It determines an optimal and the simplest approach. However, such an approach should lead to 2 important questions: (1) are these distributed 4 magnitudes correlated to some other characteristics of the measurements? (2) Is there any additional useful information hidden in these distributed results?

In order to examine the correlation between some new parameters obtained from the raw data of different measurements on 1 eye, we propose to perform at least 6 consecutive measurements. As was shown in this paper, the high correlations between values of both applanation pressures P_1 and P_2 and the respective time-dependent parameter smm make it possible to define some new parameters. These parameters are the slopes a_{p1} , a_{p2} , and a_{pa} of 3 lines, the fitting applanation pressure values P_1 , P_2 , and their arithmetic average P_a versus smm , and the respective difference Δa as well as the y-interceptions b_{p1} , b_{p2} , and b_{pa} and the difference Δb . Interestingly, the very high correlation between P_a and smm is always higher than both the high correlations between P_1 , P_2 , and smm . On the other hand, the correlations between newly introduced parameters and a_{p2} are almost always higher than those for the a_{p1} and a_{pa} parameters. Unexpectedly, lines fitting dependencies P_1 and P_2 versus smm can be parallel ($\Delta a \approx 0$), divergent ($\Delta a > 0$) or convergent ($\Delta a < 0$). It is likely that the value Δa describes some new biomechanical feature of the anterior part of the eye. However, there is no correlation between Δa and the average value of CH for measured patients. Similarly, like in Fung [16] who introduced two independent constants A and α for description of nonlinearity of the tissue elasticity, its viscoelasticity can be also described by 2 or more independent constants, not only by a single CH parameter. The proposed parameter Δa can be likely treated as an independent parameter describing corneal viscoelasticity.

The analysis of the proposed parameters for the 2 groups of patients showed that, by means of these parameters, it is possible to differentiate the group of healthy subjects from the group of glaucoma patients.

The statistical analysis carried out in this study showed that the obtained linear approximation of averaged applanation pressures P_1 and P_2 , calculated as P_a with IOPg values given by the ORA, is described by different equations (equations 9 and 10) than the one given by the ORA's producer.

The new dependencies were obtained from the raw data given directly by the ORA. Access to the raw ORA data enables a new approach to quantitative analysis of the corneal biomechanics and determination of the new parameters, in order to describe and examine some new physical features of the anterior eye. Such an approach may broaden the usefulness of the ORA instrument in ophthalmic diagnosis.

Acknowledgments

The authors are grateful for the medical assistance provided by Patrycja Krzyzanowska-Berkowska, MD from the Department of Ophthalmology, Wrocław Medical University, Poland.

Disclosure Statement

The authors report no conflicts of interest. They alone are responsible for the content and writing of the paper.

References

- 1 Luce DA: Determining in vivo biomechanical properties of the cornea with an ocular response analyzer. *J Cataract Refract Surg* 2005;31:156–162.
- 2 Reinstein DZ, Gobbe M, Archer TJ: Ocular biomechanics: measurement parameters and terminology. *J Refract Surg* 2011;27:396–397.
- 3 Kasprzak H, Mazur E, Widlicka M: Measurement and analysis of the air pressure curve on the rigid lenses by use of Ocular Response Analyzer. *Acta Bioeng Biomech* 2014;16:117–121.
- 4 Luz A, Fontes BM, Lopes B, Ramos I, Schor P, Ambrósio R: ORA waveform-derived biomechanical parameters to distinguish normal from keratoconic eyes. *Arq Bras Oftalmol* 2013;76:111–117.
- 5 Lam AKC, Chen D, Tse J: The usefulness of waveform score from the ocular response analyzer. *Optom Vis Sci* 2010;87:195–199.
- 6 Touboul D, Benard A, Mahmoud AM, Gallois A, Colin J, Roberts CJ: Early biomechanical keratoconus pattern measured with an Ocular Response Analyzer: curve analysis. *J Cataract Refract Surg* 2011;37:2144–2150.
- 7 Kotecha A, Elsheikh A, Roberts CR, Zhu H, Garway-Heath DF: Corneal thickness- and age-related biomechanical properties of the cornea measured with the Ocular Response Analyzer. *Invest Ophthalmol Vis Sci* 2006;47:5337–5347.
- 8 Lau W, Pye D: A clinical description of Ocular Response Analyzer measurements. *Invest Ophthalmol Vis Sci* 2011;52:2911–2916.
- 9 El-Malah MA: Evaluation of corneal biomechanics using ocular response analyzer for normal and primary open angle glaucoma eyes. *J Egypt Ophthalmol Soc* 2013;106:249–252.
- 10 Kotecha A, Russell RA, Sinapis A, Pourjavan S, Sinapis D, Garway-Heath DF: Biomechanical parameters of the cornea measured with the Ocular Response Analyzer in normal eyes. *BMC Ophthalmol* 2014;14:11.
- 11 Vantomme M, Pourjavan S, Detry-Morel M: The range of the waveform score of the ocular response analyzer (ORA) in healthy subjects. *Bull Soc Belge Ophthalmol* 2013;322:91–97.
- 12 Rogowska ME, Iskander DR, Kasprzak HT: Assessing subject-related variations of the Ocular Response Analyzer parameter calculation. *Clin Exp Optom* 2015;98:348–352.
- 13 Zareei A, Razeghinejad MR, Nowroozzadeh MH, Mehrabi Y, Aghazadeh-Amiri M: Intraocular pressure measurement by three different tonometers in primary congenital glaucoma. *J Ophthalmic Vis Res* 2015;10:43–48.
- 14 Bayoumi NHL, Bessa AS, El Massry AAK: Ocular Response Analyzer and Goldmann applanation tonometry: a comparative study of findings. *J Glaucoma* 2010;19:627–631.
- 15 Asejczyk-Widlicka M, Krzyżanowska-Berkowska P, Kowalska M, Iskander DR: Clinical utility of spectral analysis of intraocular pressure pulse wave. *BMC Ophthalmol* 2014;14:30.
- 16 Fung YC: *Biomechanics: Mechanical Properties of Living Tissues*. New York, Springer, 1993.