Safety and Utility of Spinal Magnetic Resonance Imaging in Patients with High-Frequency Spinal Cord Stimulators: A Prospective Single-Centre Study

James Manfield^a Richard Bartlett^b Nicholas Park^a

^aDepartment of Neurosurgery, Royal Preston Hospital, Lancashire Teaching Hospitals NHS Foundation Trust, Preston, UK; ^bDepartment of Radiology, Royal Preston Hospital, Lancashire Teaching Hospitals NHS Foundation Trust, Preston, UK

Keywords

Spinal magnetic resonance imaging · Spinal cord stimulation · High-frequency spinal cord stimulator implantation

Abstract

Background: Implanted high-frequency spinal cord stimulators at 10 kHz (HF-SCS) have recently acquired conditional approval for magnetic resonance imaging (MRI), including retrospective application to previously implanted devices. Under certain conditions, there are greater specific absorption rate (SAR) scanning restrictions compared to some conventional alternatives. This poses technical challenges to obtain diagnostic quality imaging. **Objectives:** To describe our experience with 9 such scans, demonstrating that safe and diagnostically useful images can be obtained despite these restrictions. **Methods:** We report a prospective single-centre series of 9 scans within a tertiary neuroscience centre, all obtained within the required SAR limit of ≤ 0.4 W/kg, and describe the scanning protocol we have developed. We further

illustrate this with 2 representative patient cases. **Results:** The imaging studies were well tolerated without complication. In all cases, the imaging quality was sufficient for the reporting neuroradiologist to answer the clinical question posed. **Conclusion:** Despite technical challenges, MRI is feasible, safe and diagnostically useful in HF-SCS-implanted patients. We would invite other centres that implant these devices to consider the development of their own scanning protocols to avoid the morbidity and inconvenience of implantation or computed tomography myelography. To our knowledge, this is the first reported study of MRI in HF-SCS-implanted patients achieving the requisite SAR limit of ≤ 0.4 W/kg.

© 2019 The Author(s)
Published by S. Karger AG, Basel

Introduction

Our centre has recently made high-frequency spinal cord stimulators at 10 kHz (HF-SCS) the implant of first choice, over conventional alternatives, following the publi-

Table 1. Lumbar T2-weighted MRI scanning parameters

Parameter	Routine imaging	With Nevro ¹ HF10
Repetition time	3,692	2,903
Time to echo	120	120
Number of excitations	2	1
Bandwidth	154	367.1
Reconstruction matrix	512	512
Sense	no	1.4
Voxel	0.9×1.209×4	0.9×1.16×3
Slice thickness	4 mm	4 mm
Gap	0.4 mm	0.3 mm
Field of view	AT 160 × FH 311 × LR 66	AT 200 × FH 371 × LR 56
Time	3.30 s	0.58 s

Example MRI scanning parameters for lumbar T2-weighted sagittal sequences (routine vs. HF-SCS devices in situ) using our institution's 1.5-T Philip Ingenia MRI Scanner (Philips Corporation, Amsterdam, the Netherlands). ¹ Nevro Corp., Redwood City, CA, USA.

cation of data demonstrating superior efficacy [1]. We have now implanted in excess of 100 HF-SCS devices (Nevro Senza systems; Nevro Corp., Redwood City, CA, USA) for various indications, predominantly failed back surgery syndrome. One traditional downside of these implants was the lack of magnetic resonance imaging (MRI) compatibility, although they have recently acquired conditional approval for full-body MRI, including retrospective application to previously implanted devices [2]. This is a significant advance given it has been estimated that around 82–85% of SCS-implanted patients will be expected to warrant one or more MRI scans within 5 years of implantation [3]. Given up to 50,000 such devices are implanted each year for various indications, this represents a sizeable clinical need [4].

In common with other implantable devices, there remain several technical MRI challenges which principally stem from specific absorption rate (SAR) scanning restrictions. For an implanted HF-SCS device, when imaging any portion of the spine (or any other structure approximately between the mouth and knee) the whole-body average SAR must not exceed 0.4 W/kg with a maximum active scan time of 30 min per study [2]. This is materially below the 2 W/kg limit permitted by other lower-frequency stimulator systems (e.g., Medtronic SureScan devices [Medtronic plc, Minneapolis, MN, USA]) and can be challenging to achieve without compromising diagnostic imaging quality. There are several reports in the literature of successful MRI scans with reduced SARs [5–7], with the lowest reported average SAR being 0.74 W/kg [7]. However, there have hitherto been no case series described achieving the limit of 0.4 W/kg required for

HF-SCS devices. We have now performed 9 scans in 8 patients with HF-SCS devices in situ. These scans have produced diagnostic quality images that have both identified and excluded significant pathological findings, which we describe here and illustrate with 2 case reports.

Methods

After developing spinal symptoms indicating imaging, 8 patients with implanted HF-SCS percutaneous leads were prospectively investigated by MRI scanning in our tertiary neuroscience centre during 2018. MRI was performed by our MRI radiographers using a particular adapted scanning protocol on our institution's Philips Ingenia 1.5-T MRI scanner (Philips Corporation, Amsterdam, the Netherlands). Our scanning protocol has evolved with experience, and we include our current parameters for sagittal T2-weighted lumbar sequences, alongside those for conventional, i.e. non-neuromodulation, scans as a comparison (Table 1). A company representative was present for the initial cases to ensure the implant's compatibility and check impedances to exclude lead integrity deficiency or disconnection. Following training, our in-house team can now perform these checks independently. A neuroradiologist and/or surgeon were also present to review the images obtained and assess the requirement for further sequences. An MRI physicist attended for some initial scans to aid the development of our reduced SAR scanning protocol.

Results

In all 9 cases (in 8 patients), the scan was well tolerated with none having to be abandoned due to patient discomfort. None of the patients developed new symptomatol-

Table 2. Patient and MRI details

Patient	Location of stimulator wires	Extent of MRI scan	Pathology identified or excluded	Therapeutic consequences (where applicable)
1	cervical	whole spine ¹	spinal cord, foraminal and cauda equina compression excluded	
2	thoracic	lumbar spine	neural compression following L ₅ /S ₁ decompression excluded	
3	thoracic	whole spine	new L ₅ /S ₁ right lateral disc extrusion	microdiscectomy
4	thoracic	whole spine	spinal cord and cauda equina compression excluded	
5	thoracic	whole spine	spinal cord and cauda equina compression excluded	
6	thoracic	whole spine	spinal cord and cauda equina compression excluded	
7	thoracic	cervical spine	multilevel cervical stenosis	Patient offered and considering decompressive surgery
8	thoracic	brain	cerebral infarction excluded	
8	thoracic	whole spine	spinal cord and radicular compression excluded; mild cervical stenosis identified	

¹ Our usual protocol for spinal MRI is to obtain T2-weighted sagittal views of the whole spine with axial images through the clinically/pathologically relevant regions.

ogy nor was any stimulator dysfunction identified after MRI scanning. An MRI was planned for a ninth patient but due to unsatisfactory impedance checks, suggesting either a lead not connected to the implantable pulse generator or a lead integrity deficiency, a decision was made not to proceed with the scan, and an alternative modality (CT myelography) was employed.

All 9 images were of sufficient quality for the reporting neuroradiologist to answer the clinical question being asked. Details of the 8 patients who underwent scanning and their imaging are summarised in Table 2 below.

Case Report 1

A 64-year-old male (patient 1) underwent cervical HF-SCS implantation for uncontrolled neuropathic left-sided arm pain in 2015. Initially, pain relief was excellent until he suffered a fall 18 months later. Following the incident, efficacy was lost which was refractory to reprogramming. The patient also developed left leg weakness and incoordination. Plain radiographs excluded any obvious lead migration or bony injury, hence an MRI scan was performed to exclude a spinal aetiology for his symptoms.

As can be seen from the above images (Fig. 1), the artefact from the epidural leads is fairly minimal, and the patency of the neural foramina as well as the cord signal can be evaluated. On this basis, significant spinal pathology as the cause of his leg symptoms could be excluded, and these did slowly improve over time.

Case Report 2

A 46-year-old male police officer (patient 3) underwent an HF-SCS implant in 2015 for failed back surgery syndrome, comprising back, leg and bilateral foot pain. He favoured the HF-SCS device as he is required to drive for his occupational duties. The patient had a good outcome, averaging 70% pain relief overall for over 2 years until he developed an exacerbation of right foot pain from 3/10 to 10/10 on the visual analogue scale, which was refractory to re-programming. An MRI scan was accordingly obtained.

The MRI scan revealed neural compression from prolapse of the L₅/S₁ disc (Fig. 2). The patient underwent an uncomplicated right-sided microdiscectomy following which he made a good recovery with symptomatic resolution.

Fig. 1. **a** Sagittal T2-weighted MRI of the cervicothoracic spine for patient 1. **b** Axial T2-weighted image at level C_{4/5}, corresponding to the yellow line on the sagittal image above. The epidural lead is apparent. **c** Anterior-posterior radiograph of the cervical spine showing the leads in situ without obvious fracture or migration.

Discussion

Until 2017, MRI in patients with HF-SCS implants was limited to head and extremity scans. Full-body MRI conditionality has now been acquired for all patients with percutaneous leads in situ, including those with previous implants. Presently, the minority of patients with surgically implanted paddle leads may only undergo head and extremity imaging [2]. In essence, this is due to concerns that the interaction of magnetic, radiofrequency and gradient fields with the ferromagnetic components of the stimulator apparatus might lead to equipment displacement, current induction or heating with consequent tissue damage [8].

In this series, 4 patients were scanned as an emergency, which prior to 2017 would have necessitated either surgical explantation of the stimulator or alternative investigation, e.g. CT myelography. The latter is an invasive procedure which carries small risks of significant complications including seizures and contrast reaction [9], does not image the spinal cord itself, entails ionising radiation expo-

sure and is not routinely available out of hours at our centre. Explantation is a surgical procedure and thus exposes the patient to considerable morbidity and inconvenience. In view of these factors, as well as its status as the gold standard imaging modality, MRI is now our first-line investigation in HF-SCS patients requiring spinal imaging.

There has inevitably been a learning curve with the development and refinement of our local scanning protocol to obtain adequate images within the requisite SAR limit, which initially required significant input from an MRI physicist. The scanning process for these patients is more labour intensive as it requires the presence of trained personnel to check lead impedances as well as a clinician (radiologist and/or spinal surgeon) to review the images in real time to guide the radiographer as to which further sequences are required. Typically, this entails evaluating the initial T2-weighted sagittal images and tailoring axial cuts to just the region(s) of pathological significance. The lengthened repetition time also means that the overall scan takes longer than usual, although cannot exceed 30 min of active scan time without a 60-min gap. As a result,

2

Fig. 2. **a** Sagittal T2-weighted MRI through the lumbar spine revealing prolapsed intervertebral disc at L₅/S₁. **b** Axial T2-weighted image at L₅/S₁ demonstrating the disc prolapse and resultant neural compression. **c** Anterior-posterior radiograph demonstrating satisfactory positioning of the spinal cord stimulation leads.

Fig. 3. **a, b** T2-weighted sagittal (**a**, whole spine) and axial (**b**, T_{10/11}) level MRI sequences of patient 2 demonstrating signal drop-out imaging artefact from epidural leads.

3

patients must be reasonably compliant and not claustrophobic to tolerate the scan. It is our practice to carefully counsel the patient about what to expect and endeavour to optimise analgesia prior to imaging.

Artefacts causing signal drop-out in close vicinity of the electrodes do occur but have never caused a diagnostic problem in our experience (Fig. 3). It remains possible that a subtle lesion in the region of the electrodes could be missed. Based on this series, we have been able to exclude spinal cord and cauda equina compression with confidence. Spinal cord signal change and neuroforaminal compromise can also be identified even adjacent to an electrode.

Conclusion

To our knowledge, we have presented the first case series of spinal MRI in patients harbouring HF-SCS at 10 kHz stimulators, with imaging performed at a SAR level (<0.4 W/kg) not previously described in the literature. This demonstrates that with a specifically adapted reduced SAR protocol, diagnostic quality images can be safely achieved and are well tolerated by appropriate patients. As a consequence, since 2017, we have not had to explant any patient's stimulator to facilitate imaging. On only one occasion has CT myelography been necessary (due to either a lead not connected to the implantable pulse generator or a lead integrity deficiency). We invite other centres that implant these HF-SCS stimulators to consider developing their

own MRI scanning protocol as it remains the gold standard investigation to assess the neuraxis and avoid the morbidity and inconvenience associated with myelography or explantation.

Statement of Ethics

As MRI scanning is FDA approved and was performed as part of routine clinical care, no specific IRB/ethics committee approval was needed. All patients consented to undergo MRI scanning. Patient consent for publication was not required since no identifiable data are included in the paper.

Disclosure Statement

Mr. Park has previously received sponsorship and consultancy fees from Nevro, and sponsorship from Abbott and Medtronic – all outside the submitted work. Mr. Manfield and Dr. Bartlett have no conflict. All authors have no personal, financial or institutional interest in any of the materials or devices described in this article.

Funding Sources

No funding was received for the study.

Author Contributions

James Manfield and Nicholas Park co-conceived the study and drafted the manuscript. Richard Bartlett proved radiological expertise, for images and scanning protocols, and reviewed the manuscript.

References

- 1 Kapural L, Yu C, Doust MW, Gliner BE, Vallejo R, Sitzman BT, et al. Novel 10-kHz high-frequency therapy (HF10 therapy) is superior to traditional low-frequency spinal cord stimulation for the treatment of chronic back and leg pain. The SENZA-RCT Randomized Controlled Trial. *Anesthesiology*. 2015;123(4):851–60.
- 2 Nevro: MRI guidelines for the SENZA and SENZA II systems [cited 2018 Oct 22]. Available from: <https://www.nevro.com/English/Physicians/manuals/default.aspx>
- 3 Desai MJ, Hargens LM, Breitenfeldt MD, Doth AH, Ryan MP, Gunnarsson C, et al. The rate of magnetic resonance imaging in patients with spinal cord stimulation. *Spine*. 2015 May;40(9):E531–7.
- 4 Mekhail NA, Cheng J, Narouze S, Kapural L, Mekhail MN, Deer T. Clinical applications of neurostimulation: forty years later. *Pain Pract*. 2010 Mar-Apr;10(2):103–12.
- 5 De Andres J, Valía JC, Cerda-Olmedo G, Quiroz C, Villanueva V, Martinez-Sanjuan V, et al. Magnetic resonance imaging in patients with spinal neurostimulation systems. *Anesthesiology*. 2007 Apr;106(4):779–86.
- 6 De Andres J, Martinez-Sanjuan V, Fabregat-Cid G, Asensio-Samper JM, Sanchis-Lopez N, Villanueva-Perez V. MRI-compatible spinal cord stimulator device and related changes in patient safety and imaging artifacts. *Pain Med*. 2014 Oct;15(10):1815–9.
- 7 Mutter UM, Bellut D, Porchet F, Schuknecht B. Spinal magnetic resonance imaging with reduced specific absorption rate in patients harbouring a spinal cord stimulation device - A single-centre prospective study analysing safety, tolerability and image quality. *Acta Neurochir (Wien)*. 2013 Dec;155(12):2327–32.
- 8 Rubino S, Adepoju A, Kumar V, Prusik J, Murphy N, Owusu-Sarpong S, et al. MRI Conditionality in patients with spinal cord stimulation devices. *Stereotact Funct Neurosurg*. 2016;94(4):254–8.
- 9 Sandow BA, Donnal JF. Myelography complications and current practice patterns. *AJR Am J Roentgenol*. 2005 Sep;185(3):768–71.