Validation of a New Transcutaneous tcPO₂/tcPCO₂ Sensor with an Optical Oxygen Measurement in Preterm Neonates

Willem van Weteringen^{a, b} Tanja van Essen^b Norani H. Gangaram-Panday^b
Tom G. Goos^{b, c} Rogier C.J. de Jonge^d Irwin K.M. Reiss^b

^aDepartment of Pediatric Surgery, Erasmus MC – Sophia Children's Hospital, University Medical Center Rotterdam, Rotterdam, The Netherlands; ^bDepartment of Pediatrics, Division of Neonatology, Erasmus MC – Sophia Children's Hospital, University Medical Center Rotterdam, Rotterdam, The Netherlands; ^cDepartment of Biomechanical Engineering, Faculty of Mechanical Engineering, Delft University of Technology, Delft, The Netherlands; ^dPediatric Intensive Care Unit, Departments of Pediatrics and Pediatric Surgery, Erasmus MC – Sophia Children's Hospital, University Medical Center Rotterdam, Rotterdam, The Netherlands

Keywords

Transcutaneous measurements · Transcutaneous oxygen · Transcutaneous carbon dioxide · Sensor · Neonate

Abstract

Introduction: Traditional transcutaneous oxygen (tcPO₂) measurements are affected by measurement drift, limiting accuracy and usability. The new potentially drift-free oxygen fluorescence quenching technique has been combined in a single sensor with conventional transcutaneous carbon dioxide (tcPCO₂) monitoring. This study aimed to validate optical tcPO₂ and conventional tcPCO₂ against arterial blood gas samples in preterm neonates and determine measurement drift. **Methods:** In this prospective observational study, during regular care, transcutaneous measurements were paired to arterial blood gases from preterm neonates aged 24–31 weeks of gestational age (GA) with an arterial catheter. Samples were included based on stability criteria and stratified for sepsis status. Agreement was assessed using the Bland-Altman analysis. Measurement drift per hour was calculated. **Results:** Sixty-eight premature neonates were included [median (interquartile range [IQR]) GA of 26 4/7 [25 3/7–27 5/7] weeks], resulting in 216 stable paired samples. Agreement of

stable samples in neonates without sepsis ($n = 38$) and with suspected sepsis ($n = 112$) was acceptable for tcPO₂ and good for tcPCO₂. However, in stable samples of neonates with sepsis ($n = 66$), tcPO₂ agreement (bias and 95% limits of agreement) was -32.6 (-97.0 to 31.8) mm Hg and tcPCO₂ agreement was 4.2 (-10.5 to 18.9) mm Hg. The median (IQR) absolute drift values were 0.058 (0.0231 – 0.1013) mm Hg/h for tcPO₂ and 0.30 (0.11 – 0.64) mm Hg/h for tcPCO₂. **Conclusion:** The accuracy of optical tcPO₂ in premature neonates was acceptable without sepsis, while electrochemically measured tcPCO₂ remained accurate under all circumstances. Measurement drift was negligible for tcPO₂ and highly acceptable for tcPCO₂.

© 2020 The Author(s)
Published by S. Karger AG, Basel

Introduction

Transcutaneous blood gas monitoring is widely used in neonatal intensive care [1, 2]. It provides a continuous, noninvasive alternative to arterial blood gas sampling for

T.v.E. and N.H.G.-P. contributed equally to this work.

measuring oxygen and carbon dioxide levels. The skin is arterialized by locally applying heat, causing an increase in microcirculatory blood flow. As a consequence, arterial values of oxygen and carbon dioxide can be estimated from values measured at the skin surface [3, 4]. The agreement of transcutaneously measured blood gas values with arterial values is, however, affected by factors that influence skin gas diffusion [5–7]. The diffusion capacity of the skin for oxygen is markedly lower than that for carbon dioxide and decreases with age, allowing transcutaneous oxygen (tcPO₂) levels to reach arterial levels only in thin neonatal skin [8–10]. Until now, transcutaneous blood gas measurement techniques have been based on an electrochemical measurement principle [11, 12]. Binding of other substances to the electrodes causes measurement drift over time, leading to additional inaccuracy [13]. While transcutaneous carbon dioxide (tcPCO₂) measurements are generally considered clinically useable, oxygen measurements suffer from inherent electrode instability and the fact that the electrochemical Clark-type electrode consumes oxygen, resulting in the underestimation of oxygen levels [14]. The main reasons for the sparse use of tcPO₂ monitoring are the labor intensity of frequent sensor repositioning and poor measurement accuracy. In previous studies, the large variation in patient populations, applied sensor temperatures, and mixed inclusion of capillary and arterial samples have led to a large spread in, often contradicting, study results [15]. Recently, a new transcutaneous sensor has been introduced, featuring an optical fluorescence quenching oxygen measurement, combined with a conventional electrochemical carbon dioxide sensor [16]. The fluorescence quenching technique measures oxygen levels through fluorescence decay of a dye that is quenched by oxygen, resulting in reduced fluorescence intensity and decay rates [17]. This study aimed to validate this new combined transcutaneous sensor in premature neonates, by assessing agreement with arterial blood gas samples and determining drift, while ensuring accurate comparison by applying criteria for measurement stability.

Materials and Methods

Study Set-Up

A prospective observational study was performed at a level III neonatal intensive care unit in the Netherlands. Neonates of 24 0/7 up to and including 31 6/7 weeks of gestational age (GA) with an arterial catheter were included upon the clinical indication for standard of care transcutaneous blood gas monitoring.

Transcutaneous Devices

Patients were provided with a SenTec OxiVenTTM Sensor (SenTec AG, Therwil, Switzerland) and a SenTec SDM-PO2 (SenTec Digital Monitor) for a minimum of 48 h or until the clinical indication ended. Sensor temperatures were set according to the department protocol: 42.0°C for extreme preterm neonates (≤ 25 weeks of GA) and 43.0°C for preterm neonates (26–31 weeks of GA). The site time and safety features were set to 2 h for extreme preterm neonates and 3 h for preterm neonates, automatically lowering the sensor temperature to 39°C after the measurement time elapsed. Calibration of the tcPCO₂ measurement was automatically required after the site time elapsed, the tcPO₂ measurement was calibrated automatically approximately every 24 h during a tcPCO₂ calibration. The SpO₂ channel was disabled in the neonatal mode of the transcutaneous monitor. Sensor membranes were changed every 30 days or earlier in case of any visible damage or recurrent calibration errors. Skin fixation adhesives and contact gel were used in accordance with manufacturer guidelines.

Parameters and Data Acquisition

TcPO₂ and tcPCO₂ measurements, quality indicators, sensor temperature, and calibration information were logged (Raspberry Pi 2 or 3 model B, Raspberry Pi Foundation, UK) at a 1 Hz rate. Arterial pressures were logged at a 1 Hz rate from the patient monitoring system (Dräger M540, Drägerwerk AG & Co., KGaA, Lübeck, Germany). Arterial blood samples were analyzed with a Radiometer ABL800 FLEX (Radiometer, Copenhagen, Denmark). Patient characteristics and arterial blood sample data were retrieved from the hospital patient information system (PDMS, Picis Clinical Solutions, Wakefield, MA, USA). The number of days between birth and the moment of blood sampling was presented as an indicator of development of the skin on the measurements. Ventilation parameters, infection laboratory parameters, and blood culture results were additionally documented.

Sample Selection

Arterial Sampling

Arterial pressure data were used to verify the arterial origin of the blood samples. For blood sampling, the connection between the arterial line and the pressure transducer was temporarily closed. The heparin lock was removed, a blood sample was taken, and the heparin lock was flushed. A clear pressure change could be observed for each of these steps, which was used to match the exact moment of blood sampling with the corresponding transcutaneous measurements (Fig. 1).

Sensor Temperature

Data pairs were excluded when blood sampling during standard care coincided with an elapsed transcutaneous measurement time.

Measurement Stability

Measurement stabilization and respiratory dynamics often lead to fluctuations in transcutaneous measurements. To guarantee a proper comparison between transcutaneous and arterial blood gas values, measurement stability criteria were applied to each sample. A parameter indicating the stability status was provided by the monitor. Additionally, around the exact moment of blood sampling, a time window of 10 min was analyzed for transcutaneous measurement stability (Fig. 1). The sample was marked

Fig. 1. Examples of arterial blood gas sample lookup for transcutaneous sample selection in 2 patients. The arterial pressure curve (**a**, **b**) and the corresponding transcutaneous measurements (**c**, **d**) were displayed for sample selection. The exact moment of blood withdrawal was identified from the arterial pressure curve: (1) heparin lock withdrawal, (2) arterial blood withdrawal, and (3) heparin lock flush. The tcPO_2 and tcPCO_2 values at the moment of

blood sampling were extracted, and measurements were classified as stable when there was a relative deviation of no more than 3.75 mm Hg within the 10 min around blood withdrawal. The examples show a stable (**c**) and unstable (**d**) transcutaneous measurement. tcPO_2 , transcutaneous oxygen; tcPCO_2 , transcutaneous carbon dioxide.

as unstable (Fig. 1d) when either tcPO_2 or tcPCO_2 data were missing from this time interval, or a relative deviation of more than 3.75 mm Hg was present in the tcPO_2 and/or tcPCO_2 values (Fig. 1c).

Sepsis Definition

The neonate's sepsis status was defined for each data pair to indicate the influence on skin microcirculation and potential blood gas diffusion (Fig. 2b).

Analyses

Demographics

Demographic data were reported as median and interquartile range (IQR) or as n (%). To study significance between patient groups in the subgroup analyses, the Kruskal-Wallis and Fisher's exact tests were used for continuous and categorical variables, respectively. The significance level of statistical tests was fixed at $\alpha = 0.05$.

Agreement and Correlation

Agreement between transcutaneous measurements and arterial blood gas samples was calculated according to Bland and Altman (A-B plot), accounting for multiple measurements per patient

[18]. Bias, calculated as mean difference, and limits of agreement (LoA), defined as the ± 1.96 SD of the mean difference, were calculated. Pearson's correlation coefficient was calculated to determine correlations. Agreement and correlation were determined over all stable samples and in a subgroup analysis of the potentially influencing factors: GA at birth, age of the skin, weight at sampling, sepsis status, and sensor temperature.

Drift

The drift per hour of tcPO_2 and tcPCO_2 was calculated for each data pair by subtracting the calibration values after and before blood sampling. Duplicate calibration values originating from multiple data pairs within 1 calibration interval were removed. Drift is presented as median (IQR) for tcPO_2 and tcPCO_2 , shown as absolute values due to the possibility of either positive or negative drift values.

Software

Statistical analyses were performed in R v3.5.3 (The R Foundation for Statistical Computing, Vienna, Austria). A sample selection program was built in LabVIEW (National Instruments, Austin, TX, USA).

Table 1. Patient and sample characteristics of included premature neonates and data pairs

Premature neonates, <i>n</i>	68
GA at birth, weeks	26 4/7 (25 3/7–27 5/7)
Birth weight, g	803 (708–1,006)
Gender, male	45 (66.2)
Delivery mode	
Cesarean section	48 (70.6)
Vaginal	20 (29.4)
Apgar	
At min 1	5 (3–7)
At min 5	8 (6–9)
At min 10	9 (8–9)
Umbilical cord pH	7.31 (7.25–7.34)
Multiple births	11 (16.2)
Sepsis during admission	42 (61.8)
Deceased during admission	27 (39.7)
Samples per patient	9 (4–13)
Sample, <i>n</i>	625
GA at sample, weeks	27 6/7 (26 5/7–29 4/7)
Days since birth	6 (3–12)
Weight at sampling, g	880 (680–1,195)
Ventilation mode	
Noninvasive ventilation	3 (0.5)
Invasive ventilation	288 (46.1)
High-frequency oscillatory ventilation	334 (53.4)
Sensor temperature, °C	
42	113 (18.1)
43	512 (81.9)
PaO ₂ , mm Hg	57.8 (48.8–68.3)
PaCO ₂ , mm Hg	48.0 (41.3–54.0)

Values are expressed as median (IQR) or *n* (%), unless otherwise indicated. GA, gestational age; IQR, interquartile range.

Results

Patient Demographics

Between November 2015 and April 2017, 68 premature neonates were included out of a total number of 380 neonatal intensive care unit admittances aged 26–31 weeks of GA. Neonates had a median GA of 26 4/7 (IQR 25 3/7–27 5/7) weeks at birth and a median birth weight of 803 (IQR 708–1006) g (Table 1). In the study population, the incidence of sepsis was considerably higher (61.8%) than the incidence in all admitted neonates with a GA of 26–31 weeks (22.1%) during the inclusion period. Of the 42 neonates with proven sepsis, there was only 1 case of early-onset sepsis.

Sample Characteristics

A total of 625 data pairs of arterial blood gas and transcutaneous measurements were collected and analyzed (Fig. 2a; Table 1), with a median sampled PaO₂ of 57.8 (IQR 48.8–68.3) mm Hg and PaCO₂ of 48.0 (IQR 41.3–54.0) mm Hg. No burns or skin irritation was observed in any of the neonates. The transcutaneous sensor was placed mostly at the thorax, abdomen, and lower extremity.

Sample Stability

The stability criteria were met by 216 (34.6%) data pairs. The Bland-Altman analysis showed a bias of tcPO₂ and PaO₂ data pairs of –19.1 (95% LoA –64.6 to 26.5) mm Hg (Pearson's *r* = 0.37) and a tcPCO₂ and PaCO₂ bias of 4.7 (95% LoA –7.8 to 17.1) mm Hg (Pearson's *r* = 0.85).

Subgroup and Sepsis Status Analysis

The effects of GA at birth, age of the skin, weight at the moment of blood sampling, sepsis status, and sensor temperature are shown in Table 2. Neonates in the sepsis group were older (GA: no sepsis 28 3/7 [26 4/7–29 6/7] weeks, suspected sepsis 27 6/7 [26 6/7–30 5/7] weeks, and sepsis 29 3/7 [27 6/7–31 3/7] weeks, *p* = 0.001) and had a higher weight (no sepsis 912 [622–1,260] g, suspected sepsis 930 [714–1,295] g, and sepsis 1,115 [839–1,690] g, *p* = 0.002) at the moment of blood gas sampling than the other groups, while this discrepancy was not present at birth. The Bland-Altman analysis of tcPO₂ and PaO₂ showed wide LoA for the sepsis group (Fig. 3). Agreement improved markedly for samples classified as suspected sepsis and no sepsis. In contrast, agreement of tcPCO₂ with PaCO₂ was minimally influenced by the sepsis status of the neonate.

Measurement Drift

Measurement drift was calculated for all unique calibrations of tcPO₂ (*n* = 327) and tcPCO₂ (*n* = 573). TcPO₂ drift was minimal, with a median (IQR) absolute drift of 0.058 (0.0231–0.1013) mm Hg/h, while tcPCO₂ drift was on par with the existing literature on tcPCO₂ sensors (median [IQR] drift of 0.30 [0.11–0.64] mm Hg/h) [19, 20].

Discussion

In this study, a new transcutaneous blood gas sensor using an optical technique for measuring oxygen was evaluated in a premature neonatal population during standard care. The primary outcome of this study is that

Fig. 2. a Flowchart of inclusion and exclusion of patients, data pairs, and unique calibration lines. **b** Flowchart of sepsis status definition. BC, blood culture; AB, antibiotic; CRP, C-reactive protein; tcPO₂, transcutaneous oxygen; tcPCO₂, transcutaneous carbon dioxide.

Table 2. Subgroup analysis of agreement

Parameter	Samples, <i>n</i>	tcPO ₂ -PaO ₂ , mm Hg			tcPCO ₂ -PaCO ₂ , mm Hg		
		bias	lower LoA-upper LoA	<i>R</i>	bias	lower LoA-upper LoA	<i>R</i>
GA at birth, weeks							
24 0/7 to 25 6/7	60	-18.6	-48.3 to 11.0	0.21	4.3	-9.6 to 18.2	0.76
26 0/7 to 27 6/7	94	-16.5	-58.1 to 25.2	0.36	4.9	-4.5 to 14.2	0.91
28 0/7 to 29 6/7	17	-19.8	-55.3 to 15.8	0.38	2.5	-4.9 to 9.9	0.95
30 0/7 to 31 6/7	45	-24.7	-95.6 to 46.1	0.49	5.5	-12.0 to 22.9	0.80
Postnatal skin age, week							
<1	132	-15.5	-61.1 to 30.2	0.50	5.0	-6.2 to 16.3	0.90
≥1	84	-24.7	-68.4 to 19.0	0.22	4.0	-10.3 to 18.4	0.78
Weight at sampling, kg							
<1	116	-15.7	-54.7 to 23.4	0.34	4.7	-6.9 to 16.3	0.86
≥1	100	-23.0	-74.6 to 28.6	0.41	4.6	-9.0 to 18.1	0.84
Sepsis status							
No sepsis	38	-16.1	-49.7 to 17.6	0.11	3.4	-3.9 to 10.7	0.95
Suspected sepsis	112	-12.1	-36.9 to 12.7	0.55	5.3	-7.1 to 17.8	0.87
Sepsis	66	-32.6	-97.0 to 31.8	0.45	4.2	-10.5 to 18.9	0.70
Sensor temperature, °C							
42	37	-19.7	-64.6 to 25.2	-0.40	7.8	-2.2 to 17.7	0.90
43	179	-18.9	-64.6 to 26.8	0.45	4.0	-8.6 to 16.6	0.84

Bias and 95% LoA in all stable samples related to GA at birth, age of the skin, weight at sampling, sepsis status, and sensor temperature. *r* represents Pearson's correlation coefficient. LOA, limits of agreement; GA, gestational age; tcPO₂, transcutaneous oxygen; tcPCO₂, transcutaneous carbon dioxide.

agreement of stable tcPO₂ values with PaO₂ is acceptable in suspected septic and nonseptic patients, while overall agreement of tcPCO₂ with PaCO₂ remains excellent even with sepsis. The current literature on tcPO₂ sensors shows similar results in terms of agreement [21, 22]. Conventional tcPO₂ sensors however apply a correction factor for, among other influences, the oxygen consumption of the Clark-type electrode or the so-called stirring effect [3, 14]. The investigated sensor provides the actual skin oxygen level, leading to negative agreement when oxygen diffusion is impaired. Most interesting in this study was the factor of sepsis, which caused poor agreement between tcPO₂ and PaO₂, potentially due to impairment of the microcirculation [23]. Contrary to tcPO₂, the agreement between tcPCO₂ and PaCO₂ was minimally affected by sepsis. The disparity can be explained by the higher skin diffusion capacity for carbon dioxide than for oxygen. The skin thickness in adults may have a more pronounced influence on the diffusion of and transport capacity for carbon dioxide. The effects of GA at birth in this study are also suggestive of this influence.

An important difficulty in studies is the sepsis status of the neonate. Sepsis is a continuously changing condition,

of which the effects on the microcirculation can precede the clinical manifestation by several days [23]. In this study, the definition of sepsis was based on hospital guidelines and current literature [24]. In the sepsis group, the older age and consequent development of the neonatal skin may also influence the diffusion capacity of oxygen and thus the agreement of tcPO₂ with PaO₂. Whenever there is a lack of agreement of tcPO₂ with PaO₂, edema, impaired perfusion, or incorrect sensor placement should be considered.

Transcutaneous monitoring was mostly used to monitor CO₂ during high-frequency oscillatory ventilation, used as rescue therapy. This resulted in a high prevalence of, mostly late-onset, sepsis (61.8%) and mortality (39.7%) during admission in this study population. Unfortunately, during routine care, sensor skin locations were not registered, potentially influencing agreement due to skin microcirculation heterogeneity [14, 25]. The investigated sensor was provided with skin adhesives that performed well even under humid incubator conditions and allowed for sensor repositioning for kangaroo care.

The methodological aim of this study was to remove the cause of highly varying results on transcutaneous

Fig. 3. Bland-Altman plot of the measured tcPO_2 (a) and tcPCO_2 (b) against arterial blood gas values for the defined sepsis status of the neonates. tcPO_2 , transcutaneous oxygen; tcPCO_2 , transcutaneous carbon dioxide.

monitoring in previous studies by improving the time-synchronized matching between transcutaneous measurements and arterial blood gas samples [15]. The inclusion and exclusion criteria for sample selection and measurement stability aim to improve reproducibility of future studies on transcutaneous monitoring. The criteria provide the possibility to distinguish the kinetics and dynamic properties of cardiorespiratory changes from those related to gas diffusion. The agreement analysis accounted for multiple measurements per patient, resulting in increased LoA when compared to other studies.

The investigated sensor provides a solution to the problem of tcPO_2 measurement drift, an inherent property of electrochemical blood gas sensors [13]. Although measurement drift is a well-known problem, few studies provide quantitative data that have been obtained during standard care. The conventional tcPCO_2 drift was comparable to previous studies [19, 20]. Improvements in transcutaneous measurement drift are able to significantly improve clinical usability, reliability, and accuracy. In the case of tcPO_2 , this might even increase competition with pulse oximetry [26–28], providing trend informa-

tion on reasonable estimates of the actual PaO₂ when calibrated to an arterial blood gas sample. The investigated sensor requires regular tcPCO₂ calibration, limiting the long-term accuracy advantage of the drift-free tcPO₂ measurement. A dedicated optical tcPO₂ sensor that would only need measures to prevent skin burns could provide a useful complement to neonatal oxygen monitoring.

Regardless of technical performance, the microcirculatory condition of the patient and optimal timing of blood sampling during a stable phase are of great influence on agreement of the measurements with arterial values. The influence of sepsis as demonstrated in this study is one of many potential patient-related factors that should be taken into account during the clinical use of transcutaneous monitoring.

Conclusion

The investigated combined optical tcPO₂ and conventional tcPCO₂ sensor provided acceptable agreement of tcPO₂ with PaO₂ in premature neonates without sepsis. In the case of proven sepsis, the LoA for tcPO₂ with PaO₂ widened remarkably. An equally interesting finding was the excellent agreement of tcPCO₂ with PaCO₂, despite sepsis in these premature neonates. Drift was negligible during tcPO₂ measurement and highly acceptable when measuring tcPCO₂ during standard care in premature neonates.

References

- 1 Rudiger M, Topfer K, Hammer H, Schmalisch G, Wauer RR. A survey of transcutaneous blood gas monitoring among European neonatal intensive care units. *BMC Pediatr*. 2005 Aug 10;5:30.
- 2 Ochiai M, Kurata H, Inoue H, Ichiyama M, Fujiyoshi J, Watabe S, et al. Transcutaneous blood gas monitoring among neonatal intensive care units in Japan. *Pediatr Int*. 2020 Feb; 62(2):169–74.
- 3 Lubbers DW. Theoretical basis of the transcutaneous blood gas measurements. *Crit Care Med*. 1981 Oct;9(10):721–33.
- 4 Lubbers DW. Theory and development of transcutaneous oxygen pressure measurement. *Int Anesthesiol Clin*. 1987 Fall;25(3): 31–65.
- 5 Versmold HT, Holzmann M, Linderkamp O, Riegel KP. Skin oxygen permeability in premature infants. *Pediatrics*. 1978 Oct;62(4): 488–91.
- 6 Versmold HT, Linderkamp O, Holzmann M, Strohacker I, Riegel KP. Limits of tcPO₂ monitoring in sick neonates: relation to blood pressure, blood volume, peripheral blood flow and acid base status. *Acta Anaesthesiol Scand Suppl*. 1978;68:88–90.
- 7 Lofgren O, Jacobson L. The influence of different electrode temperatures on the recorded transcutaneous PO₂ level. *Pediatrics*. 1979 Dec;64(6):892–7.
- 8 Huch R, Lübbers DW, Huch A. Quantitative continuous measurement of partial oxygen pressure on the skin of adults and new-born babies. *Pflugers Arch*. 1972;337(3):185–98.
- 9 Hansen TN, Sonoda Y, McIlroy MB. Transfer of oxygen, nitrogen, and carbon dioxide through normal adult human skin. *J Appl Physiol Respir Environ Exerc Physiol*. 1980 Sep;49(3):438–43.
- 10 Wimberley PD, Grønlund Pedersen K, Olsson J, Siggaard-Andersen O. Transcutaneous carbon dioxide and oxygen tension measured at different temperatures in healthy adults. *Clin Chem*. 1985 Oct;31(10):1611–5.
- 11 Clark LC Jr. Monitor and control of blood and tissue oxygen tensions. *ASAIO J*. 1956;2(1): 41–8.
- 12 Severinghaus JW. A combined transcutaneous PO₂-PCO₂ electrode with electrochemical HCO₃-stabilization. *J Appl Physiol Respir Environ Exerc Physiol*. 1981 Oct;51(4):1027–32.
- 13 Eberhard P, Mindt W, Jann F, Hammacher K. Continuous pO₂ monitoring in the neonate by skin electrodes. *Med Biol Eng*. 1975 May; 13(3):436–42.
- 14 Severinghaus JW, Stafford M, Thunstrom AM. Estimation of skin metabolism and blood flow with tcPO₂ and tcPCO₂ electrodes by cuff occlusion of the circulation. *Acta Anaesthesiol Scand Suppl*. 1978;68:9–15.

Acknowledgement

The authors would like to thank the nursing staff at Erasmus MC – Sophia Children's Hospital for their support of this study.

Statement of Ethics

The Medical Ethical Board of Erasmus University Medical Center Rotterdam approved the study protocol (MEC-2015-514). A waiver of informed consent was given as all data were collected as part of regular clinical care.

Conflict of Interest Statement

The authors have no conflicts of interest to declare.

Funding Sources

This study was in part funded by SenTec AG.

Author Contributions

W.W., T.G.G., R.C.J.J., and I.K.M.R. designed and set up the study. W.W. and T.E. acquired the data. W.W., T.E., and N.H.G.-P. analyzed the data. W.W. and N.H.G.-P. wrote the manuscript. T.E. and N.H.G.-P. contributed equally to this work. All authors provided input on the final manuscript.

- 15 Conway A, Tipton E, Liu WH, Conway Z, Soalheira K, Sutherland J, et al. Accuracy and precision of transcutaneous carbon dioxide monitoring: a systematic review and meta-analysis. *Thorax*. 2019 Feb;74(2):157–63.
- 16 van Weteringen W, Goos TG, van Essen T, Ellenberger C, Hayoz J, de Jonge RCJ, et al. Novel transcutaneous sensor combining optical tcPO₂ and electrochemical tcPCO₂ monitoring with reflectance pulse oximetry. *Med Biol Eng Comput*. 2020 Feb;58(2):239–47.
- 17 Opitz N, Lübbers DW. Theory and development of fluorescence-based optochemical oxygen sensors: oxygen optodes. *Int Anesthesiol Clin*. 1987 Fall;25(3):177–97.
- 18 Bland JM, Altman DG. Agreement between methods of measurement with multiple observations per individual. *J Biopharm Stat*. 2007;17(4):571–82.
- 19 Storre JH, Magnet FS, Dreher M, Windisch W. Transcutaneous monitoring as a replacement for arterial PCO₂ monitoring during nocturnal non-invasive ventilation. *Respir Med*. 2011 Jan;105(1):143–50.
- 20 Aarrestad S, Tollefsen E, Kleiven AL, Qvarfort M, Janssens JP, Skjønberg OH. Validity of transcutaneous PCO₂ in monitoring chronic hypoventilation treated with non-invasive ventilation. *Respir Med*. 2016 Mar;112:112–8.
- 21 Palmisano BW, Severinghaus JW. Transcutaneous PCO₂ and PO₂: a multicenter study of accuracy. *J Clin Monit*. 1990 Jul;6(3):189–95.
- 22 Sandberg KL, Brynjarsson H, Hjalmarson O. Transcutaneous blood gas monitoring during neonatal intensive care. *Acta Paediatr*. 2011 May;100(5):676–9.
- 23 Lush CW, Kvietys PR. Microvascular dysfunction in sepsis. *Microcirculation*. 2000 Apr;7(2):83–101.
- 24 Wynn JL, Wong HR, Shanley TP, Bizzarro MJ, Saiman L, Polin RA. Time for a neonatal-specific consensus definition for sepsis. *Pediatr Crit Care Med*. 2014 Jul;15(6):523–8.
- 25 Faltstie-Jensen N, Spaun E, Bröchner-Mortensen J, Faltstie-Jensen S. The influence of epidermal thickness on transcutaneous oxygen pressure measurements in normal persons. *Scand J Clin Lab Invest*. 1988 Oct;48(6):519–23.
- 26 Poets CF, Bassler D. Providing stability in oxygenation for preterm infants: is transcutaneous oxygen monitoring really better than pulse oximetry? *Arch Dis Child Fetal Neonatal Ed*. 2008 Sep;93(5):F330–1.
- 27 Quine D, Stenson BJ. Does the monitoring method influence stability of oxygenation in preterm infants? A randomised crossover study of saturation versus transcutaneous monitoring. *Arch Dis Child Fetal Neonatal Ed*. 2008 Sep;93(5):F347–50.
- 28 Poets CF. Noninvasive monitoring and assessment of oxygenation in infants. *Clin Perinatol*. 2019 Sep;46(3):417–33.